

P001_樹林焚化廠CEMS系統日報表-A

報表日期: 112年01月14日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	9.85	5.98	53.87	2.29	19.58	11.51	1.41	73144.62
1	10.61	1.74	52.83	2.15	19.47	11.51	1.42	73067.27
2	10.46	2.83	52.49	2.44	19.01	11.54	1.42	73846.26
3	10.49	2.23	53.91	2.41	18.75	11.60	1.41	73809.24
4	10.49	3.02	53.20	2.50	18.63	11.71	1.41	72814.29
5	11.04	4.35	52.05	2.52	18.01	11.82	1.41	71300.35
6	12.32	1.52	56.00	2.20	19.63	11.24	1.38	75515.54
7	9.44	4.07	47.54	2.52	17.49	11.83	1.37	72648.57
8	10.25	4.83	55.79	2.45	18.37	11.75	1.38	72851.93
9	9.46	2.28	50.06	2.44	18.87	11.59	1.42	73153.61
10	9.16	1.43	60.88	2.24	19.07	11.50	1.41	73607.30
11	11.11	2.88	47.74	2.45	18.42	11.72	1.41	72944.46
12	12.55	5.03	57.53	2.29	18.56	11.71	1.44	72248.00
13	7.43	6.43	50.36	2.36	17.95	12.09	1.43	70181.33
14	5.52	2.65	55.47	2.19	18.04	11.97	1.44	71426.98
15	10.98	11.83	50.62	2.64	17.12	11.98	1.44	71835.91
16	11.67	1.52	54.62	2.53	17.83	11.91	1.45	71762.30
17	14.21	5.07	45.97	2.72	16.91	11.94	1.44	71037.50
18	9.47	1.93	58.17	2.27	18.51	11.94	1.41	71350.57
19	9.72	1.38	48.84	2.31	18.81	11.79	1.38	71251.20
20	7.87	2.45	49.95	2.14	19.03	11.71	1.37	72382.24
21	10.95	5.13	53.13	2.35	18.65	11.75	1.38	72184.74
22	12.90	3.49	50.29	2.73	17.80	11.93	1.37	71918.57
23	8.94	1.78	51.00	2.50	18.07	12.15	1.40	69917.22
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	10.29	3.58	52.60	2.40	18.44	11.76	1.41	72341.67
最大值	14.21	11.83	60.88	2.73	19.63	12.15	1.47	75515.54
發生時間	17	15	10	22	6	23		6
最小值	5.52	1.38	45.97	2.14	16.91	11.24	1.36	69917.22
發生時間	14	19	17	20	17	6		23
零點偏移(率)	-0.02 %	-0.01 %	-0.01 %	-0.00 %		-0.03	0.00	0.00 %
全幅偏移(率)	3.14 %	0.18 %	-0.58 %	0.43 %		0.05	-0.03	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002_樹林焚化廠CEMS系統日報表-A

報表日期: 112年01月14日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	11.69	5.32	54.10	2.70	16.30	10.67	1.37	63002.09
1	9.95	1.47	54.15	1.98	18.90	10.38	1.38	63727.48
2	11.03	5.21	46.00	2.91	16.97	10.29	1.38	62415.30
3	10.48	4.51	49.05	2.73	16.47	10.49	1.38	62240.82
4	10.50	4.32	50.62	2.65	16.54	10.45	1.37	62692.19
5	11.15	3.69	49.31	2.56	17.08	10.23	1.36	63713.34
6	6.33	5.40	52.82	2.10	16.45	10.44	1.36	62415.33
7	10.30	2.01	50.80	2.44	16.81	10.49	1.36	62897.88
8	12.23	0.97	55.43	2.39	17.72	10.43	1.37	65060.54
9	10.06	1.73	51.36	2.53	16.02	10.84	1.40	63585.88
10	12.98	1.21	68.01	3.36	17.02	10.82	1.37	65012.62
11	15.87	9.07	49.32	4.89	15.44	11.09	1.38	63287.23
12	12.19	7.15	56.62	3.20	15.48	11.05	1.36	61733.03
13	11.17	0.67	50.64	2.55	17.32	10.57	1.36	63153.18
14	10.47	2.56	52.60	3.13	16.84	10.49	1.36	63177.36
15	10.95	1.53	57.07	3.21	16.91	10.58	1.34	63041.39
16	8.74	6.64	58.83	2.36	16.59	10.87	1.35	60993.09
17	9.85	0.69	61.49	1.79	18.55	10.50	1.37	62615.68
18	10.40	1.80	38.67	1.95	17.95	10.29	1.38	63721.26
19	10.27	1.23	49.65	1.74	18.83	10.24	1.38	64579.58
20	11.30	2.36	51.50	2.13	18.19	10.35	1.37	63989.21
21	10.50	4.60	52.30	2.44	16.39	10.79	1.36	63237.38
22	11.06	1.79	56.82	2.35	17.05	10.78	1.35	63379.33
23	10.27	2.58	48.00	2.30	17.39	10.57	1.34	62470.87
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	10.82	3.27	52.72	2.60	17.05	10.57	1.37	63172.59
最大值	15.87	9.07	68.01	4.89	18.90	11.09	1.41	65060.54
發生時間	11	11	10	11	1	11		8
最小值	6.33	0.67	38.67	1.74	15.44	10.23	1.33	60993.09
發生時間	6	13	18	19	11	5		16
零點偏移(率)	-0.24 %	-0.01 %	-0.01 %	-0.04 %		-0.02	0.00	0.00 %
全幅偏移(率)	-0.30 %	-0.01 %	-0.13 %	0.24 %		-0.02	0.00	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003_樹林焚化廠CEMS系統日報表-A

報表日期: 112年01月14日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	2.90	1.99	53.89	2.24	14.99	10.83	1.74	79741.91
1	1.41	7.84	56.15	2.05	16.23	10.72	1.74	81404.70
2	11.11	2.08	56.75	2.61	15.08	10.91	1.74	81367.91
3	11.75	1.70	57.00	2.65	15.62	10.95	1.75	82443.81
4	10.60	2.88	55.14	2.56	15.77	10.84	1.73	84093.88
5	18.74	2.77	48.73	3.02	15.33	10.79	1.72	81576.26
6	9.54	2.84	55.52	2.89	14.62	10.96	1.73	80671.02
7	8.91	1.67	52.36	2.49	15.95	10.64	1.74	74698.30
8	13.24	1.84	56.80	2.54	16.75	10.53	1.75	84466.48
9	12.86	2.80	55.94	2.51	16.64	10.47	1.77	85772.63
10	16.23	3.53	52.40	3.31	15.91	10.61	1.77	83542.32
11	14.52	3.74	53.59	3.28	16.10	10.58	1.78	83433.80
12	11.55	4.37	58.53	4.13	14.59	10.89	1.77	79898.43
13	13.30	1.11	54.85	3.73	15.69	10.74	1.76	81609.86
14	9.99	3.44	59.37	4.49	15.05	10.84	1.75	78564.65
15	10.93	3.50	52.74	4.68	15.34	10.74	1.76	80270.06
16	11.72	1.94	56.03	3.77	15.19	10.82	1.74	79551.26
17	11.41	4.09	57.24	2.52	15.76	10.70	1.75	80073.90
18	14.76	4.76	55.53	2.70	15.72	10.84	1.77	79752.06
19	10.20	3.92	55.74	2.15	16.76	10.59	1.73	79740.60
20	8.62	8.15	56.04	2.63	14.83	10.99	1.74	77999.40
21	14.65	2.29	52.32	3.02	15.13	11.08	1.75	79546.97
22	13.23	1.82	56.09	3.03	15.63	10.84	1.74	82611.83
23	12.72	2.27	52.88	3.33	15.72	10.68	1.73	81107.33
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	11.45	3.22	55.07	3.01	15.60	10.77	1.75	80997.47
最大值	18.74	8.15	59.37	4.68	16.76	11.08	1.80	85772.63
發生時間	5	20	14	15	19	21		9
最小值	1.41	1.11	48.73	2.05	14.59	10.47	1.72	74698.30
發生時間	1	13	5	1	12	9		7
零點偏移(率)	-0.08 %	-0.00 %	-0.01 %	-0.03 %		0.00	0.00	0.00 %
全幅偏移(率)	0.05 %	0.03 %	-0.34 %	-0.03 %		-0.06	-0.03	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								