

P001_樹林焚化廠CEMS系統日報表-A

報表日期: 112年03月29日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	2.20	5.12	48.95	2.58	15.45	12.67	1.46	67312.34
1	4.75	3.81	57.01	1.88	16.90	12.03	1.46	70906.21
2	2.35	4.47	71.00	2.16	16.43	13.08	1.49	63071.43
3	0.81	4.07	65.82	1.69	17.71	12.46	1.46	66147.94
4	2.59	4.72	54.15	2.12	16.70	12.24	1.46	68673.83
5	9.63	5.39	58.36	3.33	15.21	12.28	1.46	76687.60
6	2.54	15.86	74.40	3.84	14.81	14.24	1.77	54694.31
7	0.14	4.76	62.70	3.84	14.51	13.55	2.17	61234.56
8	0.16	0.77	36.72	3.06	14.96	12.59	1.64	68484.94
9	0.27	1.70	55.59	2.33	16.79	12.21	1.46	69789.74
10	7.36	1.97	48.01	2.90	15.62	12.06	1.42	71375.90
11	10.05	2.44	55.51	3.44	15.63	12.22	1.41	70237.37
12	9.92	3.40	65.57	2.80	16.38	12.64	1.44	66277.97
13	12.46	2.30	49.37	3.53	16.22	12.03	1.42	71560.28
14	9.81	2.85	43.78	3.95	15.86	12.07	1.41	71689.27
15	10.78	4.59	56.17	4.15	14.91	12.55	1.41	68141.66
16	7.72	3.48	69.78	3.17	15.73	13.14	1.42	62963.82
17	7.27	2.48	49.91	2.75	16.37	12.41	1.46	68744.36
18	11.81	1.90	47.35	2.80	16.73	11.83	1.46	72172.76
19	10.07	1.59	57.91	3.05	16.06	11.89	1.46	72300.55
20	13.99	2.23	51.40	3.40	15.77	12.01	1.46	70881.09
21	10.18	3.07	43.24	3.89	15.99	12.00	1.46	71247.28
22	13.07	3.79	51.30	3.75	15.50	12.21	1.46	70226.46
23	9.65	3.70	59.17	3.85	15.38	12.82	1.46	65550.11
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	7.07	3.77	55.55	3.09	15.90	12.47	1.50	68348.82
最大值	13.99	15.86	74.40	4.15	17.71	14.24	2.39	76687.60
發生時間	20	6	6	15	3	6		5
最小值	0.14	0.77	36.72	1.69	14.51	11.83	1.41	54694.31
發生時間	7	8	8	3	7	18		6
零點偏移(率)	-0.05 %	-0.00 %	-0.02 %	-0.00 %		-0.03	0.00	0.00 %
全幅偏移(率)	-0.15 %	0.85 %	-0.88 %	0.05 %		0.21	0.00	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002_樹林焚化廠CEMS系統日報表-A

報表日期: 112年03月29日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.75	5.23	38.14	1.55	18.61	9.98	1.45	86233.93
1	8.60	2.44	46.08	1.62	18.60	10.57	1.45	83325.32
2	8.71	3.86	45.78	1.36	19.33	10.20	1.45	85897.24
3	7.53	4.57	48.19	1.17	19.96	10.76	1.43	80920.25
4	8.10	5.57	50.02	1.82	19.57	10.60	1.42	82193.17
5	9.64	8.88	39.83	2.44	19.06	10.34	1.40	84209.66
6	9.28	12.25	49.36	4.78	20.15	9.60	1.45	82200.58
7	5.46	5.27	37.24	1.95	17.90	10.11	1.45	86248.67
8	10.38	7.05	46.87	2.16	18.61	9.98	1.44	85707.28
9	9.70	5.71	46.56	2.43	17.83	10.23	1.43	81193.51
10	12.02	7.08	55.30	2.79	17.36	10.51	1.45	80734.92
11	16.91	7.02	62.42	3.82	17.22	10.95	1.48	82023.34
12	9.61	4.47	48.68	3.48	16.59	10.59	1.49	81741.26
13	11.62	9.34	49.98	2.65	18.01	10.51	1.49	81097.94
14	12.96	12.52	51.44	4.85	17.84	10.52	1.48	81826.15
15	5.59	5.77	44.78	3.32	18.22	10.17	1.44	81747.25
16	6.38	3.02	38.39	2.04	17.13	10.29	1.46	80599.44
17	11.24	1.82	56.50	2.03	20.37	9.54	1.44	86089.34
18	9.11	2.55	43.15	2.17	17.99	10.23	1.45	83432.16
19	11.12	4.92	37.12	2.99	18.41	10.06	1.44	81780.79
20	10.38	1.38	41.48	2.90	18.83	9.95	1.45	82608.19
21	11.49	2.44	42.13	2.92	17.62	10.46	1.44	81091.84
22	9.22	3.07	38.69	2.77	19.09	10.02	1.44	82155.19
23	10.33	2.66	38.19	2.17	19.48	9.88	1.45	84113.27
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	9.71	5.37	45.68	2.59	18.49	10.25	1.45	82882.11
最大值	16.91	12.52	62.42	4.85	20.37	10.95	1.49	86248.67
發生時間	11	14	11	14	17	11		7
最小值	5.46	1.38	37.12	1.17	16.59	9.54	1.39	80599.44
發生時間	7	20	19	3	12	17		16
零點偏移(率)	-0.03 %	-0.05 %	-0.04 %	0.00 %		-0.06	0.00	0.00 %
全幅偏移(率)	0.60 %	0.81 %	-0.99 %	0.24 %		0.01	0.00	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003_樹林焚化廠CEMS系統日報表-A

報表日期: 112年03月29日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	4.78	4.31	57.82	1.33	17.88	10.65	1.43	84073.39
1	5.76	2.07	60.44	1.43	18.65	10.43	1.45	89754.41
2	13.04	5.96	48.04	1.56	17.99	10.14	1.43	91216.23
3	4.91	2.78	60.88	1.06	20.32	10.48	1.43	87677.58
4	9.03	2.54	64.37	1.55	19.13	10.87	1.44	85977.96
5	3.65	2.10	54.09	1.45	19.90	10.84	1.47	85633.31
6	5.38	8.31	50.36	2.45	20.69	9.57	1.48	94988.86
7	5.74	4.35	53.83	1.60	16.96	10.81	1.43	92906.89
8	8.02	7.33	52.55	2.21	15.99	10.70	1.43	87636.45
9	6.98	7.16	52.16	4.40	14.39	11.13	1.43	83162.50
10	10.27	3.92	59.00	3.27	14.30	11.28	1.43	82504.72
11	16.33	2.85	57.34	4.50	14.56	11.18	1.42	80503.16
12	12.96	1.64	53.33	4.90	14.19	11.70	1.43	80982.56
13	11.29	2.75	49.24	2.74	16.71	10.51	1.41	78623.61
14	11.40	7.40	47.63	3.36	16.87	10.33	1.40	79919.72
15	11.52	8.59	51.00	5.69	14.38	10.87	1.43	78562.11
16	10.42	5.07	49.23	3.24	15.40	10.62	1.43	79436.79
17	10.20	8.83	45.08	2.15	17.36	10.15	1.43	83316.54
18	9.40	4.03	44.38	2.56	16.62	10.29	1.43	82510.34
19	10.96	1.67	51.70	2.31	16.81	10.50	1.43	80574.85
20	11.47	4.25	50.06	2.62	16.72	10.51	1.43	82785.33
21	10.03	3.10	52.89	2.37	16.38	10.69	1.43	80941.46
22	13.24	3.17	53.02	2.71	16.45	10.83	1.43	80335.58
23	9.93	1.88	59.17	2.33	18.29	10.35	1.43	85759.16
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	9.45	4.42	53.23	2.66	16.96	10.64	1.43	84157.65
最大值	16.33	8.83	64.37	5.69	20.69	11.70	1.49	94988.86
發生時間	11	17	4	15	6	12		6
最小值	3.65	1.64	44.38	1.06	14.19	9.57	1.38	78562.11
發生時間	5	12	18	3	12	6		15
零點偏移(率)	-0.06 %	-0.00 %	-0.02 %	-0.00 %		-0.02	0.00	0.00 %
全幅偏移(率)	1.64 %	0.85 %	-1.07 %	0.11 %		0.18	-0.03	0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								