

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 113年01月05日

報表日期: 113年01月05日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.63	9.60	64.05	0.57	17.22	11.49	1.02	82786.57
1	9.27	9.19	47.00	0.87	17.46	10.87	1.02	87103.62
2	9.61	8.86	44.21	1.11	18.13	11.07	1.04	84865.13
3	10.01	10.02	55.06	2.20	15.27	11.69	1.06	83121.44
4	9.20	8.77	53.83	2.30	15.58	11.51	1.06	84461.77
5	10.45	4.57	50.89	1.59	16.48	11.17	1.03	81367.60
6	10.65	8.17	53.99	2.51	14.49	11.93	1.02	81279.83
7	14.94	5.24	56.05	10.05	12.77	12.44	1.02	78590.34
8	9.49	5.53	57.19	1.89	14.74	12.00	1.04	80665.13
9	8.84	4.19	55.63	1.52	15.11	11.73	1.05	82956.60
10	9.14	4.87	53.62	1.92	14.75	11.98	1.03	80623.14
11	8.16	5.41	62.72	0.90	15.98	12.13	1.06	78094.94
12	9.39	6.16	43.49	1.02	16.07	11.45	1.05	83729.62
13	9.69	6.20	59.03	0.75	16.86	11.71	1.03	80752.36
14	8.93	5.74	55.38	1.16	16.75	11.34	1.03	84238.55
15	10.04	5.93	65.94	0.97	16.66	11.77	1.04	80948.33
16	8.17	5.74	52.84	0.85	16.33	11.47	1.07	83933.66
17	11.02	7.22	47.80	2.10	14.92	11.53	1.07	84743.27
18	8.84	10.08	47.07	2.92	14.00	11.84	1.06	82439.87
19	12.32	9.65	49.14	2.89	13.49	12.01	1.05	81374.72
20	8.39	4.07	50.22	3.20	12.93	12.60	1.04	76987.30
21	9.09	10.13	51.53	2.14	13.97	12.12	1.03	80614.39
22	9.75	9.38	54.43	1.38	15.05	11.66	1.02	83585.43
23	8.38	9.69	56.55	1.89	14.49	12.07	1.02	80669.69
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	9.64	7.27	53.65	2.03	15.40	11.73	1.04	82080.55
最 大 值	14.94	10.13	65.94	10.05	18.13	12.60	1.09	87103.62
發生時間	7	21	15	7	2	20		1
最 小 值	7.63	4.07	43.49	0.57	12.77	10.87	1.00	76987.30
發生時間	0	20	12	0	7	1		20
零點偏移(率)	-0.27 %	-0.00 %	-0.13 %	-0.03 %		-0.04		0.20 %
全幅偏移(率)	0.16 %	1.23 %	0.21 %	0.28 %		-0.10		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

報表日期: 113年01月05日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.87	10.46	55.68	1.64	15.16	11.14	1.02	79410.53
1	6.37	9.49	51.41	1.53	17.18	10.80	1.02	80462.29
2	9.24	8.72	37.15	2.56	17.24	10.28	1.02	84685.96
3	9.44	9.07	40.21	3.57	14.35	10.95	1.03	81605.63
4	6.51	7.75	46.70	2.85	15.16	11.01	1.05	81861.92
5	6.21	7.86	42.32	3.05	14.33	10.87	1.06	82927.06
6	1.37	29.90	42.29	3.45	13.93	10.81	1.04	83235.73
7	7.72	3.53	48.44	4.76	13.73	11.13	1.04	83149.32
8	9.99	7.46	53.58	2.23	13.26	11.37	1.05	81235.47
9	7.62	3.16	54.58	2.70	13.47	11.75	1.01	81088.69
10	6.97	6.92	50.87	2.44	15.11	11.46	1.04	82912.35
11	5.60	6.59	51.51	2.24	15.39	11.47	1.00	82611.90
12	7.69	7.89	45.65	2.95	14.80	11.40	0.99	83966.08
13	4.63	8.35	44.92	2.83	14.60	11.51	1.00	83054.74
14	5.88	6.12	45.64	3.31	14.51	11.39	1.01	84091.22
15	1.89	6.06	50.89	2.72	13.53	12.17	0.99	77848.85
16	2.94	5.63	50.34	3.20	13.94	11.45	0.98	82925.92
17	0.65	1.57	51.56	2.75	13.58	11.63	1.02	82642.98
18	2.18	4.03	47.95	2.71	14.16	11.10	1.03	84574.34
19	5.58	12.41	47.40	2.88	13.67	11.24	1.04	81689.07
20	7.39	4.32	49.10	3.96	12.24	11.38	1.04	79304.31
21	3.71	5.04	48.15	2.98	14.07	11.26	1.04	80966.63
22	8.47	3.58	51.07	2.67	15.42	10.60	1.02	84872.15
23	7.11	3.80	49.95	1.79	16.73	10.45	1.02	84894.16
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	5.96	7.49	48.22	2.82	14.57	11.19	1.02	82334.05
最 大 值	9.99	29.90	55.68	4.76	17.24	12.17	1.06	84894.16
發生時間	8	6	0	7	2	15		23
最 小 值	0.65	1.57	37.15	1.53	12.24	10.28	0.97	77848.85
發生時間	17	17	2	1	20	2		15
零點偏移(率)	-0.04 %	-0.01 %	-0.04 %	-0.04 %		0.00		0.10 %
全幅偏移(率)	-0.46 %	0.11 %	0.35 %	0.01 %		0.08		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 113年01月05日

項目 單位 小時\限標	#3-HCl_C ppm	#3-CO_C ppm	#3-NOx_C ppm	#3-SO2_C ppm	#3-H2O %	#3-O2_C %	#3-OPAC %	#3-FLOW_C m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	11.24	5.02	58.46	2.22	16.24	10.23	1.04	77342.11
1	9.52	2.76	47.83	0.88	17.94	9.33	1.04	81284.30
2	6.01	4.17	38.89	0.92	16.85	9.57	1.04	78185.61
3	13.47	3.44	41.65	2.55	14.44	10.26	1.04	73758.48
4	15.83	5.41	37.90	2.95	15.97	9.86	1.04	76146.16
5	12.16	0.88	37.21	2.47	16.04	9.91	1.04	76566.44
6	12.34	2.41	37.74	4.32	15.73	9.65	1.05	78882.40
7	6.73	5.52	40.85	2.99	17.89	9.39	1.09	75096.78
8	6.85	4.66	40.56	1.45	17.16	9.66	1.09	79107.40
9	8.92	3.61	53.34	1.58	14.25	10.95	1.09	73037.71
10	3.30	5.24	59.32	1.30	13.87	12.44	1.09	67041.63
11	8.67	2.53	55.22	1.49	15.04	11.15	1.10	77056.64
12	7.38	8.63	49.12	2.09	13.30	11.51	1.10	74771.56
13	1.21	9.02	42.61	1.72	13.14	11.08	1.10	73665.43
14	0.84	2.80	50.25	0.50	16.37	10.28	1.09	78629.96
15	1.52	6.22	46.75	0.67	15.90	10.27	1.10	78703.75
16	2.15	7.26	47.38	0.72	15.28	10.42	1.09	75955.19
17	0.93	5.56	47.22	1.56	13.44	10.63	1.09	73363.60
18	4.04	7.59	46.62	2.30	13.12	10.57	1.09	73303.33
19	8.83	5.73	47.71	2.41	13.86	10.54	1.07	75279.82
20	7.20	7.72	54.25	2.64	11.22	11.54	1.08	72649.69
21	8.09	6.37	56.54	2.58	12.07	11.46	1.09	76748.92
22	10.70	4.74	53.83	3.45	12.54	11.51	1.08	77928.81
23	3.18	5.23	57.73	1.42	14.26	11.19	1.06	74763.99
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	7.13	5.11	47.87	1.97	14.83	10.56	1.07	75802.90
最 大 值	15.83	9.02	59.32	4.32	17.94	12.44	1.11	81284.30
發生時間	4	13	10	6	1	10		1
最 小 值	0.84	0.88	37.21	0.50	11.22	9.33	1.02	67041.63
發生時間	14	5	5	14	20	1		10
零點偏移(率)	-0.06 %	-0.00 %	-0.07 %	-0.02 %		0.02		0.40 %
全幅偏移(率)	0.00 %	-0.28 %	-0.09 %	0.03 %		0.22		0.40 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								