

P001_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月14日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.96	12.18	54.85	0.38	18.44	11.10	0.99	82781.68
1	8.05	12.86	55.39	0.22	22.12	10.34	0.95	85214.98
2	7.15	13.52	58.42	1.01	17.40	11.64	0.96	79014.25
3	5.35	11.79	65.17	0.23	19.67	10.88	0.98	84560.87
4	6.86	12.24	64.74	0.34	19.29	11.13	0.99	81668.55
5	1.84	10.64	62.41	0.23	19.76	11.04	1.01	83429.12
6	4.43	12.02	54.61	0.20	20.25	10.32	1.00	87526.79
7	8.53	11.19	49.95	0.46	18.37	10.88	0.99	85263.12
8	8.50	3.97	55.44	0.69	18.18	10.99	0.99	83906.48
9	7.92	4.25	50.34	1.02	16.90	11.27	1.02	82615.31
10	8.39	4.94	52.23	1.03	16.74	11.25	0.99	79884.82
11	7.14	4.59	55.56	0.57	17.11	11.22	1.01	82033.91
12	4.98	4.18	65.09	0.35	18.68	11.08	1.01	82220.59
13	7.18	4.22	52.23	0.32	18.55	10.98	0.99	82929.65
14	7.90	4.67	61.47	0.58	17.71	11.58	0.99	78864.01
15	11.00	4.28	54.37	0.71	18.38	10.80	0.99	84241.12
16	7.02	6.93	45.48	0.89	16.66	11.20	0.99	79441.73
17	5.65	3.88	55.21	0.58	17.32	11.35	0.99	79692.30
18	4.83	3.92	56.46	0.27	18.74	11.09	1.01	82947.72
19	10.14	4.46	52.88	0.46	18.18	10.91	1.01	84467.88
20	10.55	8.39	51.78	0.84	17.45	11.05	0.99	83703.47
21	8.72	10.99	53.60	0.51	17.27	11.27	0.99	82740.42
22	8.74	10.69	56.43	0.41	17.77	11.19	0.99	82777.24
23	9.34	12.02	57.41	0.38	17.89	11.13	0.99	82416.02
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	7.42	8.03	55.90	0.53	18.28	11.07	0.99	82680.92
最大值	11.00	13.52	65.17	1.03	22.12	11.64	1.03	87526.79
發生時間	15	2	3	10	1	2		6
最小值	1.84	3.88	45.48	0.20	16.66	10.32	0.94	78864.01
發生時間	5	17	16	6	16	6		14
零點偏移(率)	-0.30 %	0.00 %	-0.01 %	-0.00 %		0.03		0.00 %
全幅偏移(率)	0.01 %	0.19 %	-0.07 %	0.69 %		-0.12		0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月14日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	4.42	7.00	59.95	0.96	16.83	10.79	1.10	77398.99
1	0.14	7.80	65.45	0.73	18.41	10.99	1.10	74886.89
2	0.38	8.45	56.77	0.55	17.60	10.57	1.10	74544.35
3	0.84	6.54	58.48	0.63	18.70	9.91	1.10	80954.64
4	0.70	7.31	57.27	0.51	18.55	10.09	1.10	81492.07
5	2.24	10.35	50.29	1.22	17.29	9.82	1.10	77081.51
6	1.01	5.20	53.64	0.36	18.11	9.81	1.10	73844.05
7	4.79	6.25	54.97	0.50	18.98	9.47	1.12	78188.71
8	5.49	3.01	44.03	0.56	19.04	9.64	1.12	76296.94
9	2.94	3.68	45.59	0.37	19.29	9.29	1.11	79568.16
10	3.56	2.73	45.19	0.54	17.78	9.59	1.13	78463.26
11	4.89	13.19	56.42	0.56	19.26	9.68	1.13	78845.04
12	3.93	2.29	55.02	0.54	18.17	9.66	1.12	77273.48
13	1.66	4.16	44.02	0.20	18.70	9.28	1.12	73439.97
14	0.99	4.51	54.18	0.26	19.39	9.12	1.12	79221.02
15	6.87	5.29	45.10	0.86	17.59	9.86	1.12	79164.75
16	7.41	2.68	55.66	1.07	16.61	10.28	1.12	81065.26
17	4.60	2.08	54.56	1.05	16.77	10.54	1.12	77716.07
18	7.24	2.13	56.49	0.97	17.47	10.22	1.12	80259.25
19	9.92	3.22	60.95	1.21	16.79	10.42	1.12	78850.96
20	9.16	5.32	53.15	0.91	17.34	10.22	1.12	80730.54
21	10.42	7.31	56.90	1.09	16.67	10.24	1.12	80080.66
22	9.49	7.10	57.93	0.94	17.06	10.23	1.12	80281.97
23	10.04	7.32	53.07	0.76	17.76	10.04	1.11	81206.66
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	4.71	5.62	53.96	0.72	17.92	9.99	1.11	78368.97
最大值	10.42	13.19	65.45	1.22	19.39	10.99	1.14	81492.07
發生時間	21	11	1	5	14	1		4
最小值	0.14	2.08	44.02	0.20	16.61	9.12	1.09	73439.97
發生時間	1	17	13	13	16	14		13
零點偏移(率)	-0.05 %	-0.00 %	-0.06 %	-0.07 %		-0.01		-0.10 %
全幅偏移(率)	0.04 %	0.51 %	-0.89 %	0.79 %		-0.05		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月14日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	2.53	9.78	56.33	1.38	17.28	10.98	1.04	81808.45
1	3.48	12.53	52.35	1.71	16.80	11.06	1.01	82637.19
2	5.97	10.64	59.99	2.37	14.95	11.30	0.99	81073.62
3	4.36	9.41	54.85	2.31	14.69	11.26	1.00	81413.46
4	6.36	10.28	58.68	2.23	15.38	11.16	1.01	81810.39
5	6.78	10.76	57.70	2.02	16.25	11.04	1.00	82957.29
6	4.97	11.31	51.43	2.34	15.30	11.00	0.99	78971.44
7	2.07	8.54	55.98	1.66	16.17	11.15	1.04	80899.41
8	3.40	4.81	51.71	1.56	16.77	10.76	1.00	82647.38
9	4.93	5.13	57.36	2.44	14.50	11.39	1.04	77226.50
10	5.97	3.33	55.15	1.74	15.97	10.95	1.00	80914.11
11	6.72	10.49	55.20	2.10	15.48	11.26	0.99	78855.87
12	6.16	3.72	54.88	1.70	16.19	11.12	1.01	78814.06
13	2.06	3.82	61.90	1.29	17.43	11.20	1.04	76992.33
14	6.69	6.70	45.93	1.56	18.08	10.41	1.04	82721.97
15	9.77	9.46	42.74	2.02	16.97	10.59	1.03	76489.32
16	11.69	5.82	47.61	2.50	15.93	10.56	1.02	78177.68
17	5.86	3.63	46.09	2.01	15.24	11.00	0.98	75659.13
18	6.66	5.92	54.32	2.13	15.43	11.20	1.00	79166.19
19	8.17	2.05	54.58	1.73	16.54	10.84	1.03	83037.59
20	9.83	11.05	56.44	1.72	16.36	10.85	1.01	82009.09
21	11.56	11.82	57.95	1.97	15.78	11.14	0.99	82285.96
22	11.10	10.45	55.20	2.19	15.44	11.21	0.98	80954.41
23	12.99	11.76	52.89	2.13	15.44	11.12	0.97	79988.85
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	6.67	8.05	54.05	1.95	16.02	11.02	1.01	80312.99
最大值	12.99	12.53	61.90	2.50	18.08	11.39	1.06	83037.59
發生時間	23	1	13	16	14	9		19
最小值	2.06	2.05	42.74	1.29	14.50	10.41	0.97	75659.13
發生時間	13	19	15	13	9	14		17
零點偏移(率)	-0.06 %	-0.03 %	-0.01 %	-0.07 %		0.01		0.00 %
全幅偏移(率)	0.04 %	0.41 %	-0.82 %	0.35 %		0.31		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								