

P001\_樹林焚化廠CEMS系統日報表-A

報表日期: 112年02月19日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	0.96	7.10	59.41	3.62	14.15	13.18	1.46	63277.30
1	0.24	7.72	68.50	3.29	14.96	13.33	1.46	61982.88
2	2.90	4.21	51.10	3.11	14.56	12.42	1.41	69611.09
3	9.35	8.90	49.42	4.42	13.19	12.52	1.40	69941.78
4	8.88	9.74	54.22	4.51	12.71	12.68	1.44	68834.15
5	2.55	4.94	62.31	3.81	13.64	12.56	1.49	66299.57
6	10.29	1.78	65.24	3.73	13.94	12.57	1.47	68869.28
7	10.76	1.95	55.02	3.87	13.79	12.56	1.46	69067.31
8	9.81	2.04	61.15	3.62	13.87	12.31	1.47	70875.46
9	10.78	5.06	58.52	3.73	13.99	12.39	1.49	70108.02
10	10.41	7.74	55.37	4.15	13.42	12.73	1.47	68300.53
11	8.23	5.50	67.94	3.52	14.54	12.80	1.51	66721.34
12	9.75	4.86	59.86	3.38	14.71	12.44	1.49	69999.53
13	10.48	4.15	65.52	4.02	14.04	12.34	1.48	71250.75
14	10.23	7.07	53.55	3.76	14.40	12.50	1.44	69714.47
15	9.79	3.60	65.70	3.55	14.27	12.44	1.48	70888.86
16	10.32	8.25	50.92	4.19	13.37	12.73	1.49	68466.66
17	10.83	4.62	61.01	3.59	14.26	12.44	1.46	70736.79
18	12.16	4.99	65.52	3.69	14.28	12.33	1.45	72282.16
19	10.51	3.25	61.21	3.68	14.06	12.44	1.44	71169.75
20	9.59	3.65	55.09	4.25	12.95	12.59	1.43	70737.86
21	8.99	3.03	62.11	3.31	14.63	12.23	1.45	72557.20
22	9.22	3.84	59.60	3.15	14.76	12.23	1.46	72372.78
23	7.86	5.82	59.94	3.35	14.67	12.21	1.45	72503.59
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	8.54	5.16	59.51	3.72	14.05	12.54	1.46	69440.38
最大值	12.16	9.74	68.50	4.51	14.96	13.33	1.57	72557.20
發生時間	18	4	1	4	1	1		21
最小值	0.24	1.78	49.42	3.11	12.71	12.21	1.39	61982.88
發生時間	1	6	3	2	4	23		1
零點偏移(率)	-0.04 %	-0.00 %	0.00 %	-0.00 %		0.00	0.00	0.00 %
全幅偏移(率)	3.39 %	0.90 %	-0.60 %	0.37 %		0.09	0.00	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002\_樹林焚化廠CEMS系統日報表-A

報表日期: 112年02月19日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.64	3.70	43.68	1.50	18.84	10.53	1.43	81390.59
1	9.49	5.02	60.29	1.99	18.51	10.45	1.43	82918.17
2	9.78	6.19	47.23	3.07	16.33	10.63	1.41	79807.08
3	10.58	5.46	49.73	2.88	17.20	10.23	1.38	82257.31
4	11.05	6.37	46.51	3.03	16.36	10.07	1.38	79079.11
5	10.42	9.43	50.32	2.49	16.89	10.15	1.40	82463.71
6	5.17	7.31	60.55	1.96	16.65	10.63	1.45	82283.89
7	10.63	3.60	59.02	2.54	16.78	10.71	1.45	81651.37
8	9.39	3.80	55.11	2.53	17.66	10.38	1.47	84686.16
9	9.23	3.64	53.73	2.17	17.03	10.77	1.43	82959.11
10	10.47	9.88	52.00	4.04	17.31	10.21	1.43	80076.49
11	10.02	5.66	56.13	2.21	18.13	9.89	1.40	81253.58
12	10.14	2.32	57.07	2.17	17.80	10.26	1.39	82696.54
13	10.33	4.11	58.41	1.85	17.91	10.38	1.44	83787.82
14	9.14	3.38	44.43	2.31	16.04	10.91	1.43	79904.32
15	10.18	2.53	46.04	2.42	16.06	10.69	1.39	80758.72
16	10.32	2.05	55.90	2.29	16.42	10.53	1.43	81846.10
17	10.53	2.21	49.90	2.92	15.66	10.63	1.46	80506.03
18	11.19	2.81	59.07	2.63	15.86	10.62	1.46	82443.65
19	15.01	2.09	57.08	3.00	16.08	10.61	1.46	84910.31
20	22.01	2.94	62.85	3.03	16.50	10.75	1.45	85457.72
21	9.98	2.49	50.05	2.22	16.50	10.48	1.45	84422.39
22	12.11	3.97	58.50	2.55	15.37	10.87	1.45	83084.88
23	10.10	4.08	56.75	2.27	16.46	10.49	1.45	83681.21
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	10.62	4.38	53.76	2.50	16.85	10.49	1.43	82263.59
最大值	22.01	9.88	62.85	4.04	18.84	10.91	1.47	85457.72
發生時間	20	10	20	10	0	14		20
最小值	5.17	2.05	43.68	1.50	15.37	9.89	1.38	79079.11
發生時間	6	16	0	0	22	11		4
零點偏移(率)	-0.04 %	-0.00 %	-0.06 %	-0.02 %		-0.04	0.00	0.00 %
全幅偏移(率)	1.61 %	1.36 %	0.06 %	0.77 %		0.15	-0.03	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003\_樹林焚化廠CEMS系統日報表-A

報表日期: 112年02月19日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	4.88	4.26	57.98	1.47	17.19	10.81	1.72	83207.20
1	4.86	6.27	62.48	1.62	16.74	11.18	1.72	81895.44
2	6.55	10.33	52.82	3.27	14.76	10.98	1.70	81782.59
3	3.24	5.16	51.18	2.71	13.31	11.59	1.70	76307.33
4	6.08	5.42	60.27	2.59	14.03	11.34	1.68	78859.71
5	7.28	3.65	58.58	2.28	15.40	11.20	1.71	82985.39
6	9.77	3.96	58.68	2.46	14.72	11.39	1.72	81776.34
7	4.46	9.07	68.37	2.13	15.75	12.78	1.74	73108.75
8	4.18	6.39	58.69	2.29	15.05	12.11	1.78	75062.88
9	10.58	3.87	54.75	2.34	15.49	11.05	1.72	81847.65
10	8.64	4.28	58.92	2.19	15.33	11.11	1.72	81225.74
11	7.50	5.29	57.02	2.12	15.96	10.92	1.73	81625.30
12	8.88	3.96	60.05	2.38	15.51	11.28	1.73	79240.26
13	7.61	3.63	57.87	2.00	16.19	10.91	1.70	79015.20
14	9.05	8.50	57.54	1.96	16.60	10.38	1.69	82142.17
15	10.50	5.55	60.71	2.64	14.34	11.41	1.69	79314.27
16	14.17	4.51	62.02	3.03	14.30	11.25	1.72	81273.89
17	11.59	3.44	62.52	2.77	14.12	11.45	1.69	80040.74
18	11.37	5.14	61.16	2.89	14.22	11.35	1.68	81318.53
19	11.71	7.13	59.77	2.90	15.81	10.52	1.67	87399.34
20	11.75	5.03	63.45	2.77	14.81	11.18	1.67	84438.52
21	9.54	3.05	63.57	2.01	15.72	11.03	1.67	84618.19
22	8.04	4.72	57.83	1.99	15.39	10.90	1.67	84562.80
23	8.50	4.25	61.01	2.22	15.37	10.87	1.67	83603.48
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	8.36	5.29	59.47	2.38	15.25	11.21	1.70	81110.49
最大值	14.17	10.33	68.37	3.27	17.19	12.78	2.02	87399.34
發生時間	16	2	7	2	0	7		19
最小值	3.24	3.05	51.18	1.47	13.31	10.38	1.67	73108.75
發生時間	3	21	3	0	3	14		7
零點偏移(率)	-0.01 %	-0.03 %	-0.12 %	-0.02 %		0.00	0.00	0.00 %
全幅偏移(率)	1.67 %	2.17 %	0.03 %	0.72 %		0.10	0.02	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								