

## P001 樹林焚化廠CEMS系統日報表-A

報表日期: 115年07月18日

| 項目<br>單位<br><br>小時\限標 | #1-HCl_C | #1-CO_C  | #1-NOx_C | #1-SO2_C | #1-H2O   | #1-O2_C  | #1-OPAC  | #1-FLOW_C |
|-----------------------|----------|----------|----------|----------|----------|----------|----------|-----------|
|                       | ppm      | ppm      | ppm      | ppm      | %        | %        | %        | m3/Hour   |
|                       | 60.00    | 150.00   | 220.00   | 150.00   | 100.00   | 25.00    | 10.00    | 999999.00 |
| 0                     | 7.56     | 12.01    | 52.69    | 2.26     | 18.50    | 10.82    | 1.26     | 84389.25  |
| 1                     | 10.52    | 12.70    | 52.98    | 2.78     | 19.59    | 10.66    | 1.26     | 84550.18  |
| 2                     | 9.03     | 12.76    | 53.62    | 1.72     | 18.89    | 10.98    | 1.25     | 82863.14  |
| 3                     | 9.25     | 12.58    | 49.21    | 2.18     | 17.47    | 10.95    | 1.25     | 84255.46  |
| 4                     | 10.44    | 12.94    | 52.01    | 2.26     | 17.39    | 11.12    | 1.25     | 82779.29  |
| 5                     | 7.05     | 11.54    | 54.91    | 2.42     | 16.70    | 11.38    | 1.23     | 79798.92  |
| 6                     | 9.12     | 12.08    | 57.22    | 3.11     | 16.84    | 11.23    | 1.24     | 82402.95  |
| 7                     | 7.79     | 12.05    | 59.29    | 1.42     | 18.81    | 11.02    | 1.23     | 82293.48  |
| 8                     | 9.24     | 11.71    | 53.19    | 1.66     | 17.89    | 11.15    | 1.22     | 82057.63  |
| 9                     | 10.00    | 11.09    | 37.40    | 1.80     | 19.59    | 10.17    | 1.22     | 88134.36  |
| 10                    | 9.05     | 10.84    | 46.40    | 1.74     | 19.05    | 10.43    | 1.22     | 86406.90  |
| 11                    | 10.65    | 11.12    | 56.37    | 2.53     | 17.71    | 10.89    | 1.22     | 83954.13  |
| 12                    | 12.60    | 11.24    | 58.12    | 2.66     | 16.81    | 11.07    | 1.22     | 83403.35  |
| 13                    | 7.53     | 11.85    | 58.68    | 1.80     | 16.51    | 11.65    | 1.20     | 78980.83  |
| 14                    | 9.48     | 11.95    | 60.41    | 2.17     | 16.92    | 11.57    | 1.20     | 79589.87  |
| 15                    | 7.10     | 12.03    | 58.65    | 1.17     | 18.67    | 11.17    | 1.20     | 81110.18  |
| 16                    | 7.91     | 11.93    | 50.69    | 1.92     | 16.78    | 11.40    | 1.20     | 81084.29  |
| 17                    | 9.66     | 11.47    | 58.48    | 2.09     | 16.79    | 11.29    | 1.21     | 81791.08  |
| 18                    | 9.60     | 11.19    | 60.19    | 1.72     | 17.51    | 11.00    | 1.21     | 83487.54  |
| 19                    | 8.92     | 10.99    | 61.98    | 1.45     | 18.67    | 10.59    | 1.21     | 85782.96  |
| 20                    | 8.93     | 12.45    | 51.62    | 1.87     | 18.59    | 10.68    | 1.22     | 84954.55  |
| 21                    | 7.17     | 12.82    | 56.58    | 1.32     | 19.10    | 10.97    | 1.23     | 81902.25  |
| 22                    | 10.65    | 12.62    | 55.96    | 1.49     | 17.81    | 11.04    | 1.21     | 82660.08  |
| 23                    | 6.89     | 13.34    | 52.58    | 1.58     | 18.82    | 11.20    | 1.22     | 80726.75  |
| 監測時數                  | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 超限次數                  | 0        | 0        | 0        | 0        | 0        | 0        | 0        | 0         |
| 有效時數                  | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 平 均 值                 | 9.01     | 11.97    | 54.55    | 1.96     | 17.98    | 11.02    | 1.22     | 82889.98  |
| 最 大 值                 | 12.60    | 13.34    | 61.98    | 3.11     | 19.59    | 11.65    | 1.29     | 88134.36  |
| 發生時間                  | 12       | 23       | 19       | 6        | 1        | 13       |          | 9         |
| 最 小 值                 | 6.89     | 10.84    | 37.40    | 1.17     | 16.51    | 10.17    | 1.19     | 78980.83  |
| 發生時間                  | 23       | 10       | 9        | 15       | 13       | 9        |          | 13        |
| 零點偏移(率)               | -0.07 %  | 0.01 %   | -0.03 %  | -0.04 %  |          | -0.03    | -0.01    | 0.10 %    |
| 全幅偏移(率)               | 0.75 %   | 1.18 %   | 0.83 %   | 0.79 %   |          | -0.22    | -0.06    | 0.10 %    |
| 有效比率                  | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 %  |
| 備註                    |          |          |          |          |          |          |          |           |

報表日期: 115年07月18日

| 項目<br>單位<br>小時\限標 | #2-HCl_C | #2-CO_C  | #2-NOx_C | #2-SO2_C | #2-H2O   | #2-O2_C  | #2-OPAC  | #2-FLOW_C |
|-------------------|----------|----------|----------|----------|----------|----------|----------|-----------|
|                   | ppm      | ppm      | ppm      | ppm      | %        | %        | %        | m3/Hour   |
|                   | 60.00    | 150.00   | 220.00   | 150.00   | 100.00   | 25.00    | 10.00    | 999999.00 |
| 0                 | 5.59     | 13.28    | 49.79    | 1.04     | 18.43    | 10.96    | 0.97     | 81994.88  |
| 1                 | 11.45    | 13.36    | 59.74    | 1.68     | 17.81    | 11.25    | 0.97     | 81000.26  |
| 2                 | 11.55    | 13.44    | 57.73    | 1.36     | 17.73    | 11.34    | 0.95     | 80240.33  |
| 3                 | 2.93     | 13.63    | 58.89    | 1.29     | 16.99    | 11.48    | 0.97     | 80015.44  |
| 4                 | 5.88     | 13.27    | 63.31    | 1.53     | 17.74    | 11.12    | 0.97     | 82393.34  |
| 5                 | 3.69     | 12.97    | 51.92    | 0.95     | 18.17    | 11.07    | 0.95     | 82249.22  |
| 6                 | 5.34     | 12.53    | 52.24    | 0.69     | 18.54    | 11.03    | 1.02     | 85913.63  |
| 7                 | 7.77     | 12.26    | 56.02    | 1.35     | 17.11    | 10.98    | 1.01     | 83732.31  |
| 8                 | 8.35     | 12.20    | 67.50    | 1.45     | 17.73    | 10.75    | 1.02     | 85204.68  |
| 9                 | 4.52     | 11.79    | 57.60    | 0.77     | 19.23    | 10.34    | 1.02     | 86472.32  |
| 10                | 3.16     | 12.18    | 55.27    | 0.83     | 18.14    | 10.82    | 1.01     | 84500.25  |
| 11                | 6.84     | 12.63    | 64.42    | 1.49     | 16.75    | 11.35    | 1.01     | 81190.29  |
| 12                | 11.67    | 12.38    | 60.48    | 1.43     | 17.08    | 11.12    | 1.01     | 81748.56  |
| 13                | 1.57     | 13.03    | 58.29    | 0.96     | 17.55    | 11.55    | 1.02     | 78673.51  |
| 14                | 9.41     | 12.38    | 52.38    | 1.13     | 17.53    | 11.06    | 1.02     | 82395.51  |
| 15                | 6.53     | 12.72    | 63.10    | 1.39     | 16.93    | 11.38    | 1.00     | 79899.18  |
| 16                | 2.55     | 12.77    | 45.56    | 1.05     | 17.45    | 11.21    | 0.98     | 81643.95  |
| 17                | 11.08    | 12.38    | 61.21    | 1.02     | 17.81    | 10.94    | 0.99     | 83256.72  |
| 18                | 7.71     | 12.42    | 62.68    | 0.61     | 19.00    | 10.64    | 0.99     | 84924.75  |
| 19                | 3.19     | 12.06    | 59.07    | 0.47     | 19.44    | 10.19    | 1.00     | 87487.59  |
| 20                | 2.71     | 13.80    | 57.59    | 0.91     | 18.59    | 10.92    | 1.01     | 82217.09  |
| 21                | 7.16     | 13.83    | 52.55    | 0.57     | 18.59    | 11.02    | 0.99     | 81906.39  |
| 22                | 7.55     | 14.02    | 61.84    | 0.67     | 18.85    | 10.84    | 0.99     | 82600.58  |
| 23                | 2.06     | 13.10    | 56.57    | 0.16     | 20.46    | 10.16    | 1.01     | 87040.94  |
| 監測時數              | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 超限次數              | 0        | 0        | 0        | 0        | 0        | 0        | 0        | 0         |
| 有效時數              | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 平 均 值             | 6.26     | 12.85    | 57.74    | 1.03     | 18.07    | 10.98    | 0.99     | 82862.57  |
| 最 大 值             | 11.67    | 14.02    | 67.50    | 1.68     | 20.46    | 11.55    | 1.05     | 87487.59  |
| 發生時間              | 12       | 22       | 8        | 1        | 23       | 13       |          | 19        |
| 最 小 值             | 1.57     | 11.79    | 45.56    | 0.16     | 16.75    | 10.16    | 0.94     | 78673.51  |
| 發生時間              | 13       | 9        | 16       | 23       | 11       | 23       |          | 13        |
| 零點偏移(率)           | -0.06 %  | -0.07 %  | -0.04 %  | -0.00 %  |          | -0.07    | -0.01    | -0.10 %   |
| 全幅偏移(率)           | -1.09 %  | 0.53 %   | 0.72 %   | 0.11 %   |          | -0.30    | -0.06    | 0.00 %    |
| 有效比率              | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 %  |
| 備註                |          |          |          |          |          |          |          |           |

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 115年07月18日

| 項目<br>單位<br>小時\限標 | #3-HCl_C | #3-CO_C  | #3-NOx_C | #3-SO2_C | #3-H2O   | #3-O2_C  | #3-OPAC  | #3-FLOW_C |
|-------------------|----------|----------|----------|----------|----------|----------|----------|-----------|
|                   | ppm      | ppm      | ppm      | ppm      | %        | %        | %        | m3/Hour   |
|                   | 60.00    | 150.00   | 220.00   | 150.00   | 100.00   | 25.00    | 10.00    | 999999.00 |
| 0                 | 7.76     | 12.47    | 49.89    | 2.45     | 20.88    | 9.27     | 1.27     | 77765.51  |
| 1                 | 7.51     | 13.57    | 43.91    | 2.12     | 21.62    | 9.11     | 1.27     | 78451.20  |
| 2                 | 5.89     | 12.13    | 54.14    | 2.82     | 19.76    | 9.38     | 1.28     | 77068.75  |
| 3                 | 7.81     | 12.37    | 56.32    | 2.76     | 19.32    | 9.66     | 1.31     | 75049.20  |
| 4                 | 5.66     | 12.11    | 52.49    | 2.64     | 18.87    | 9.51     | 1.29     | 77017.17  |
| 5                 | 8.44     | 12.51    | 58.07    | 3.03     | 18.96    | 9.87     | 1.27     | 75049.78  |
| 6                 | 9.26     | 11.06    | 54.44    | 2.96     | 20.13    | 9.36     | 1.30     | 79159.54  |
| 7                 | 3.58     | 11.01    | 42.23    | 2.39     | 19.67    | 9.37     | 1.29     | 80042.36  |
| 8                 | 8.12     | 10.99    | 52.08    | 2.71     | 18.27    | 9.59     | 1.31     | 78163.95  |
| 9                 | 5.76     | 11.19    | 62.82    | 2.53     | 18.09    | 9.96     | 1.27     | 76327.95  |
| 10                | 10.17    | 11.09    | 58.50    | 2.91     | 17.38    | 9.84     | 1.28     | 77261.64  |
| 11                | 9.19     | 11.23    | 61.98    | 2.77     | 17.82    | 10.01    | 1.31     | 75954.53  |
| 12                | 8.56     | 11.52    | 58.98    | 2.48     | 16.65    | 10.46    | 1.31     | 72799.02  |
| 13                | 7.73     | 11.73    | 60.46    | 2.42     | 17.05    | 10.55    | 1.29     | 74190.87  |
| 14                | 9.12     | 12.49    | 62.01    | 4.01     | 15.76    | 10.58    | 1.27     | 73538.11  |
| 15                | 1.32     | 12.61    | 60.05    | 2.56     | 14.93    | 10.65    | 1.28     | 73965.49  |
| 16                | 1.57     | 11.90    | 56.79    | 2.11     | 16.51    | 10.50    | 1.30     | 73925.94  |
| 17                | 5.59     | 13.46    | 61.42    | 2.23     | 17.77    | 10.34    | 1.28     | 76550.19  |
| 18                | 3.19     | 13.36    | 66.59    | 2.27     | 17.83    | 10.91    | 1.27     | 73653.54  |
| 19                | 3.19     | 14.03    | 63.06    | 2.21     | 18.46    | 10.67    | 1.28     | 75261.60  |
| 20                | 9.32     | 12.94    | 59.71    | 2.68     | 17.86    | 10.38    | 1.31     | 77222.14  |
| 21                | 9.15     | 12.96    | 62.41    | 2.64     | 17.90    | 10.40    | 1.31     | 77111.48  |
| 22                | 4.56     | 12.82    | 59.30    | 2.26     | 17.93    | 10.31    | 1.30     | 78494.44  |
| 23                | 6.50     | 14.57    | 62.48    | 2.75     | 17.43    | 11.32    | 1.28     | 71093.02  |
| 監測時數              | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 超限次數              | 0        | 0        | 0        | 0        | 0        | 0        | 0        | 0         |
| 有效時數              | 24       | 24       | 24       | 24       | 24       | 24       | 24       | 24        |
| 平 均 值             | 6.62     | 12.34    | 57.51    | 2.61     | 18.20    | 10.08    | 1.29     | 76046.56  |
| 最 大 值             | 10.17    | 14.57    | 66.59    | 4.01     | 21.62    | 11.32    | 1.34     | 80042.36  |
| 發生時間              | 10       | 23       | 18       | 14       | 1        | 23       |          | 7         |
| 最 小 值             | 1.32     | 10.99    | 42.23    | 2.11     | 14.93    | 9.11     | 1.26     | 71093.02  |
| 發生時間              | 15       | 8        | 7        | 16       | 15       | 1        |          | 23        |
| 零點偏移(率)           | -0.03 %  | -0.00 %  | -0.04 %  | 0.00 %   |          | -0.07    | -0.01    | 0.20 %    |
| 全幅偏移(率)           | 1.50 %   | 0.20 %   | 1.38 %   | 1.72 %   |          | 0.26     | 0.00     | 0.20 %    |
| 有效比率              | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 % | 100.00 %  |
| 備註                |          |          |          |          |          |          |          |           |