

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 114年08月12日

報表日期: 114年08月12日

[illegible]

報表日期: 114年08月12日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.95	9.90	51.16	2.09	17.36	10.85	0.99	80640.57
1	8.55	8.31	56.58	2.41	17.47	10.89	1.00	79892.44
2	9.36	7.98	60.27	1.94	17.99	10.75	1.00	80509.51
3	8.28	7.94	57.65	1.45	18.03	10.72	1.01	80497.61
4	9.32	7.85	57.87	1.33	18.45	10.53	1.01	83082.14
5	8.99	8.10	58.31	2.02	17.02	10.99	1.01	80291.15
6	2.95	7.92	63.51	0.80	17.89	10.79	1.02	87129.95
7	7.95	8.65	59.41	1.80	18.04	11.42	1.02	76470.61
8	8.71	8.22	59.36	1.90	18.10	10.92	1.02	80148.49
9	9.28	9.60	58.53	1.92	18.56	10.47	1.02	83750.89
10	8.96	8.66	54.03	2.28	18.04	10.60	1.02	83221.92
11	4.89	10.97	54.12	2.83	16.25	10.97	1.01	82086.85
12	8.80	10.50	58.20	3.06	15.79	10.97	1.01	80202.67
13	9.59	9.04	61.38	2.12	15.58	10.64	1.00	79061.51
14	8.13	7.30	53.26	2.87	15.55	10.74	0.99	78190.80
15	9.73	9.14	52.04	2.57	16.50	10.68	0.99	79335.94
16	10.39	8.73	56.93	3.05	15.52	11.14	1.00	80899.13
17	8.25	11.05	63.17	3.46	16.51	11.10	1.02	80759.13
18	8.03	8.18	59.89	2.47	17.89	10.74	1.02	82030.06
19	9.28	8.63	57.20	2.58	17.37	10.85	1.02	81826.12
20	9.16	9.49	55.11	3.66	15.73	10.88	1.01	82243.20
21	9.39	10.46	54.43	3.34	15.98	10.97	1.01	83129.78
22	10.01	8.97	59.93	3.28	16.09	11.11	1.01	80719.36
23	7.25	8.30	57.43	1.70	17.76	10.93	1.01	81335.23
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	8.51	8.91	57.49	2.37	17.06	10.86	1.01	81143.96
最 大 值	10.39	11.05	63.51	3.66	18.56	11.42	1.03	87129.95
發生時間	16	17	6	20	9	7		6
最 小 值	2.95	7.30	51.16	0.80	15.52	10.47	0.99	76470.61
發生時間	6	14	0	6	16	9		7
零點偏移(率)	-0.19 %	-0.01 %	-0.01 %	-0.03 %		-0.08		-0.10 %
全幅偏移(率)	0.95 %	-0.01 %	0.31 %	0.05 %		-0.01		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 114年08月12日

項目 單位 小時\限標	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	9.17	7.83	56.31	1.86	18.39	9.62	1.41	78985.47
1	7.44	9.17	58.10	1.85	18.01	10.00	1.41	76934.58
2	9.15	7.39	57.18	2.13	17.90	9.70	1.42	77217.65
3	9.69	8.92	58.77	1.95	17.57	9.70	1.42	78949.68
4	8.30	7.49	56.19	1.84	17.05	9.99	1.43	77317.23
5	8.38	8.04	58.50	1.74	17.34	9.84	1.44	76974.09
6	9.34	8.85	56.21	2.19	16.27	10.07	1.45	76479.02
7	8.16	9.05	54.63	2.25	16.04	10.04	1.40	75270.70
8	8.77	7.86	65.74	2.08	16.49	10.33	1.42	77110.09
9	7.60	7.57	61.65	2.09	16.77	10.19	1.38	77359.56
10	7.22	8.19	59.03	1.84	17.10	10.02	1.40	79274.78
11	2.22	8.88	56.17	2.05	15.54	10.27	1.42	75902.28
12	4.28	9.58	60.61	1.83	17.05	9.92	1.40	75846.89
13	6.99	11.35	47.69	3.57	13.26	10.22	1.36	68788.29
14	8.10	9.93	48.02	1.49	14.61	9.86	1.36	68635.69
15	10.65	11.68	54.67	2.50	15.83	9.71	1.36	70452.20
16	8.51	11.03	56.67	2.24	16.15	9.49	1.37	74869.77
17	7.85	7.32	62.93	1.91	17.79	9.53	1.38	76379.20
18	8.81	7.21	61.14	2.17	19.18	9.27	1.38	81136.93
19	8.58	7.69	50.50	1.85	18.25	9.56	1.41	82934.11
20	9.79	7.78	51.80	1.86	17.44	9.82	1.40	82416.17
21	8.08	8.18	49.23	2.06	16.24	10.16	1.39	80535.21
22	9.78	8.84	57.31	2.48	15.71	10.25	1.37	75464.91
23	8.83	6.24	60.93	2.25	16.69	10.31	1.38	73635.28
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	8.15	8.59	56.67	2.09	16.78	9.91	1.40	76619.57
最 大 值	10.65	11.68	65.74	3.57	19.18	10.33	1.47	82934.11
發生時間	15	15	8	13	18	8		19
最 小 值	2.22	6.24	47.69	1.49	13.26	9.27	1.36	68635.69
發生時間	11	23	13	14	13	18		14
零點偏移(率)	-0.11 %	-0.07 %	-0.06 %	-0.01 %		-0.25		0.20 %
全幅偏移(率)	-0.77 %	0.09 %	-0.14 %	0.28 %		0.11		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								