

P001 樹林焚化廠CEMS系統日報表-A

報表日期: 113年02月14日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	3.78	8.61	56.38	0.39	17.58	11.34	1.01	82996.01
1	4.64	9.16	41.75	0.49	16.47	11.50	1.01	82213.21
2	5.84	10.07	57.75	1.49	15.55	12.20	1.03	77600.65
3	3.73	14.15	57.21	1.38	14.46	13.58	1.07	66316.69
4	3.12	11.61	51.96	0.84	15.12	13.09	1.10	70918.26
5	2.72	7.13	56.77	0.58	16.64	12.06	1.07	72678.77
6	3.60	8.49	45.51	0.66	16.20	11.90	1.01	80148.11
7	6.23	8.77	43.40	1.88	14.01	11.72	0.99	83528.93
8	10.00	9.14	48.01	1.93	14.58	11.85	1.01	82087.12
9	9.26	7.83	47.67	1.79	14.84	11.92	1.00	80891.55
10	8.91	7.54	49.79	1.06	16.10	11.49	1.01	83133.05
11	5.01	7.37	48.96	0.99	16.25	11.61	1.01	81844.91
12	4.27	10.15	42.99	1.41	15.44	11.60	1.00	82694.81
13	9.03	4.53	50.88	0.58	16.94	10.99	1.01	86445.12
14	10.17	0.45	51.60	0.96	16.19	11.17	1.03	85427.12
15	7.33	0.45	58.23	0.51	16.46	11.26	1.04	84250.77
16	9.52	3.62	50.29	1.50	14.84	11.59	1.03	83259.44
17	10.34	8.99	43.66	2.08	14.20	11.62	1.00	83796.79
18	9.34	5.88	51.93	1.53	15.43	11.68	1.01	81811.75
19	9.92	6.13	45.00	1.64	14.40	11.99	1.02	79749.56
20	9.34	5.19	58.82	0.67	16.03	11.66	1.00	81973.93
21	8.63	5.53	40.63	0.89	16.18	11.66	1.01	81645.01
22	6.80	6.57	54.57	0.37	17.13	11.61	1.03	81351.06
23	9.36	7.19	43.94	1.01	16.15	11.63	1.01	82037.01
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	7.12	7.27	49.90	1.11	15.72	11.78	1.02	80783.32
最 大 值	10.34	14.15	58.82	2.08	17.58	13.58	1.18	86445.12
發生時間	17	3	20	17	0	3		13
最 小 值	2.72	0.45	40.63	0.37	14.01	10.99	0.97	66316.69
發生時間	5	14	21	22	7	13		3
零點偏移(率)	-0.02 %	-0.04 %	-0.01 %	0.00 %		-0.02		0.20 %
全幅偏移(率)	0.02 %	-0.21 %	-0.09 %	0.01 %		0.02		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002 樹林焚化廠CEMS系統日報表-A

報表日期: 113年02月14日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	2.16	5.46	44.75	1.06	16.08	10.58	1.03	80401.00
1	2.44	6.10	42.79	1.33	15.80	10.96	1.05	80018.72
2	2.22	5.75	47.50	1.89	15.24	11.72	1.08	77994.66
3	2.07	6.55	58.10	1.19	16.25	11.66	1.07	77691.10
4	1.70	6.69	51.09	0.76	17.27	11.13	1.07	80473.60
5	1.50	6.46	40.92	0.80	17.87	10.64	1.06	83499.99
6	3.33	6.80	62.22	0.87	16.11	12.03	1.04	78958.99
7	3.12	6.79	36.64	1.59	15.69	11.07	1.04	82466.89
8	2.41	5.99	43.58	2.26	13.61	11.97	1.06	76042.36
9	7.65	2.94	49.35	1.84	15.62	11.36	1.04	79515.76
10	6.17	2.92	45.63	1.97	15.28	11.56	1.05	78084.11
11	3.61	6.05	47.71	1.41	16.01	11.22	1.04	79410.34
12	3.64	2.90	47.97	1.98	15.26	11.56	1.04	77704.24
13	3.04	2.59	44.52	2.01	15.21	11.37	1.05	78802.00
14	3.37	2.78	41.79	2.07	16.06	10.94	1.05	81432.72
15	7.10	2.81	41.98	2.78	14.73	11.34	1.04	79679.32
16	5.19	3.35	39.80	2.41	15.36	11.03	1.04	82038.24
17	4.86	3.36	42.30	2.43	14.87	11.23	1.02	80600.87
18	9.45	3.29	41.40	2.08	14.88	11.29	1.03	80411.61
19	8.64	3.55	38.32	2.18	15.01	11.21	1.06	80214.67
20	5.72	3.36	40.51	1.61	15.85	11.09	1.03	80988.23
21	2.33	3.67	44.08	1.73	14.98	11.54	1.02	77946.68
22	4.74	3.78	41.70	1.88	15.50	11.31	1.03	79357.50
23	7.19	3.56	40.11	1.97	15.69	11.10	1.06	80984.02
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	4.32	4.48	44.78	1.75	15.59	11.29	1.05	79779.90
最 大 值	9.45	6.80	62.22	2.78	17.87	12.03	1.10	83499.99
發生時間	18	6	6	15	5	6		5
最 小 值	1.50	2.59	36.64	0.76	13.61	10.58	1.02	76042.36
發生時間	5	13	7	4	8	0		8
零點偏移(率)	-0.05 %	0.00 %	-0.03 %	-0.03 %		0.06		0.00 %
全幅偏移(率)	0.02 %	-0.70 %	-0.80 %	0.12 %		0.13		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

