

P001_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月29日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.59	10.59	49.76	1.47	17.85	11.58	1.04	80787.94
1	9.50	10.56	52.75	1.88	17.32	11.84	1.05	77953.40
2	8.18	10.52	55.20	1.71	17.86	11.51	1.05	76890.01
3	7.09	10.50	46.84	2.31	15.87	12.07	1.05	75716.13
4	9.38	8.73	51.62	1.74	17.18	11.77	1.05	77881.42
5	11.34	13.25	53.25	2.84	16.03	11.92	1.05	76809.26
6	9.49	10.69	53.99	2.19	17.01	11.88	1.05	77132.02
7	9.61	10.48	49.87	2.13	17.33	11.39	1.05	80294.30
8	9.20	10.16	51.85	1.87	16.90	11.69	1.04	78087.43
9	9.45	13.29	50.21	1.96	16.35	11.65	1.02	78264.82
10	9.47	14.95	47.52	3.72	14.09	12.13	1.05	74242.80
11	9.11	10.60	48.12	2.74	14.87	12.02	1.05	76310.47
12	9.65	9.92	58.46	1.57	17.31	11.70	1.03	78014.03
13	9.87	13.89	50.67	2.42	15.81	11.88	1.02	77250.27
14	9.18	12.80	39.08	2.56	15.79	11.75	1.05	76914.29
15	9.22	10.72	49.69	2.04	16.26	11.79	1.04	76950.35
16	4.47	5.46	67.67	0.98	17.30	11.76	1.04	77316.18
17	7.81	4.85	36.87	1.50	17.60	11.56	1.03	79166.71
18	9.21	6.16	66.85	0.81	19.35	10.99	1.05	82124.08
19	8.79	5.79	47.51	1.61	17.15	11.42	1.04	80939.83
20	8.85	7.30	55.40	1.42	17.81	11.39	1.05	80593.37
21	5.64	8.68	57.36	1.21	17.46	11.90	1.05	76468.19
22	9.14	8.91	55.39	1.39	18.25	11.16	1.03	81823.29
23	6.34	9.41	61.90	1.10	18.30	11.79	1.02	77158.85
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	8.69	9.93	52.41	1.88	16.96	11.69	1.04	78128.73
最大值	11.34	14.95	67.67	3.72	19.35	12.13	1.06	82124.08
發生時間	5	10	16	10	18	10		18
最小值	4.47	4.85	36.87	0.81	14.09	10.99	1.01	74242.80
發生時間	16	17	17	18	10	18		10
零點偏移(率)	-0.70 %	-0.10 %	-0.02 %	-0.01 %		0.00		0.00 %
全幅偏移(率)	0.02 %	0.34 %	-0.03 %	0.03 %		0.02		0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月29日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.53	9.49	57.18	2.19	16.92	10.75	1.20	79074.56
1	9.47	4.45	51.19	2.16	18.28	9.98	1.20	83163.95
2	9.83	5.74	61.98	1.53	18.15	9.74	1.20	83694.34
3	9.20	15.64	50.18	1.98	18.57	9.94	1.18	84815.79
4	9.23	10.65	53.90	2.07	17.97	10.12	1.19	84444.50
5	9.78	4.19	57.95	2.77	17.24	10.35	1.18	82648.94
6	10.70	7.70	52.79	1.86	18.93	9.50	1.18	88133.19
7	8.84	5.81	56.60	2.54	17.25	10.32	1.18	82684.80
8	9.23	9.27	50.26	2.19	17.49	9.98	1.18	81548.36
9	10.63	7.87	49.40	2.15	17.99	9.63	1.18	81901.79
10	8.71	12.01	49.10	1.66	17.64	9.37	1.18	76218.92
11	9.37	11.25	52.63	1.61	17.49	9.15	1.18	77245.19
12	9.81	10.51	59.60	1.17	17.42	8.97	1.18	79811.64
13	8.26	11.10	58.59	1.66	16.25	8.84	1.18	75210.81
14	9.35	13.28	53.43	2.12	15.86	8.85	1.19	72031.89
15	10.12	8.41	57.55	1.89	16.13	8.96	1.18	75294.45
16	9.22	11.54	33.73	3.89	17.64	8.57	1.18	78287.07
17	9.33	11.60	41.81	2.40	16.33	9.34	1.18	80732.14
18	8.67	6.17	51.36	3.01	14.91	10.69	1.18	81526.09
19	9.45	3.45	47.30	2.22	17.52	10.23	1.18	83496.21
20	7.83	4.99	59.70	1.31	17.89	10.22	1.18	83674.68
21	6.59	5.89	59.85	1.72	16.87	10.91	1.18	78650.04
22	5.38	6.62	66.17	0.66	19.03	10.27	1.17	82761.40
23	8.39	6.21	53.49	1.03	19.02	10.16	1.18	83455.52
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	8.96	8.49	53.57	1.99	17.45	9.78	1.18	80854.43
最大值	10.70	15.64	66.17	3.89	19.03	10.91	1.20	88133.19
發生時間	6	3	22	16	22	21		6
最小值	5.38	3.45	33.73	0.66	14.91	8.57	1.15	72031.89
發生時間	22	19	16	22	18	16		14
零點偏移(率)	-0.02 %	-0.06 %	-0.01 %	-0.00 %		0.00		0.00 %
全幅偏移(率)	0.04 %	0.29 %	-0.44 %	0.46 %		-0.22		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003_樹林焚化廠CEMS系統日報表-A

報表日期: 112年07月29日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	0.51	14.95	46.76	1.10	15.11	10.93	0.99	75654.99
1	8.54	12.24	61.30	2.25	14.51	11.54	1.01	75006.69
2	11.97	5.21	66.47	1.47	16.69	11.08	1.02	82010.28
3	9.34	14.32	61.24	1.79	15.26	11.36	1.02	81516.13
4	14.01	16.03	58.01	2.76	15.14	11.27	1.02	80604.68
5	11.21	5.48	61.25	2.05	14.98	11.32	1.02	78253.81
6	7.50	11.06	59.66	1.65	15.44	11.21	1.02	80123.40
7	7.20	9.85	55.38	1.58	15.89	10.98	1.01	75218.90
8	9.40	12.02	52.66	1.88	15.71	10.84	0.99	78854.08
9	12.14	10.34	52.01	2.52	14.99	11.38	0.99	80026.67
10	9.03	11.41	56.46	2.59	14.49	11.47	1.02	76247.36
11	7.76	9.64	50.77	2.05	14.43	10.68	0.97	73793.51
12	10.10	10.24	51.15	1.87	15.07	10.46	1.00	75984.92
13	8.46	9.04	43.73	2.59	13.21	10.53	1.01	71580.44
14	12.17	8.13	49.16	3.14	13.05	10.92	1.02	71468.92
15	7.89	8.25	49.53	1.56	15.29	10.38	1.02	75776.45
16	5.89	6.44	49.18	2.34	13.22	10.85	0.98	69990.85
17	0.31	9.13	61.91	1.05	15.18	11.12	0.97	68819.72
18	0.00	6.34	51.77	1.28	14.55	11.22	0.97	69584.25
19	0.01	8.07	57.94	1.44	14.22	11.45	0.99	68295.52
20	0.00	7.60	58.93	0.87	16.05	11.14	0.99	74185.12
21	0.06	7.17	57.53	0.78	16.64	10.59	1.02	78341.22
22	0.18	10.32	62.00	1.06	15.34	11.64	1.01	75905.39
23	2.60	6.83	63.77	0.62	16.38	11.62	0.97	71764.34
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	6.51	9.59	55.77	1.76	15.04	11.08	1.00	75375.32
最大值	14.01	16.03	66.47	3.14	16.69	11.64	1.06	82010.28
發生時間	4	4	2	14	2	22		2
最小值	0.00	5.21	43.73	0.62	13.05	10.38	0.97	68295.52
發生時間	18	2	13	23	14	15		19
零點偏移(率)	-0.04 %	-0.01 %	-0.04 %	-0.00 %		-0.04		-0.10 %
全幅偏移(率)	-0.49 %	0.06 %	-0.64 %	0.67 %		0.15		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								