

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 113年02月25日

報表日期: 113年02月25日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.32	6.81	56.66	1.28	16.36	11.53	1.02	83294.46
1	9.49	6.62	51.48	1.14	17.49	11.33	1.04	84116.71
2	9.05	6.52	59.85	0.92	18.01	11.13	1.05	85460.43
3	9.20	6.46	48.79	0.93	17.65	11.25	1.04	84921.94
4	9.32	6.43	36.92	1.57	16.61	11.37	1.03	84716.67
5	10.91	13.13	69.21	1.31	15.81	11.92	1.02	81679.28
6	6.25	14.38	44.60	0.87	16.04	12.46	1.02	75841.67
7	6.67	9.21	43.52	0.93	17.00	11.65	1.01	82302.49
8	6.36	6.20	54.72	0.84	16.31	12.05	0.99	79052.09
9	4.50	6.46	60.74	0.80	16.32	12.37	0.99	76043.58
10	8.15	6.46	53.39	0.94	16.60	11.89	0.99	80257.16
11	9.47	7.26	54.30	1.07	16.41	11.93	0.99	80134.03
12	9.22	7.80	48.96	1.28	16.18	11.70	0.99	81988.60
13	9.94	5.86	51.51	1.72	15.87	11.58	1.00	83595.28
14	9.76	5.94	56.04	1.29	16.61	11.41	1.00	84252.87
15	9.81	5.94	46.41	2.22	14.74	11.86	0.99	82280.39
16	10.39	6.11	53.71	2.34	14.46	11.85	1.02	82394.51
17	9.80	6.04	61.86	1.37	15.73	11.59	1.02	83336.23
18	9.52	6.58	54.74	2.32	14.31	12.02	1.03	81463.85
19	9.86	5.53	58.19	1.29	16.28	11.53	1.02	83919.88
20	9.29	6.19	55.42	1.50	15.45	11.66	1.02	83344.44
21	8.66	6.26	54.72	1.58	15.10	11.85	1.02	82104.75
22	9.55	5.02	62.42	1.67	14.67	11.84	1.02	82393.77
23	9.36	5.93	47.49	1.92	14.96	11.85	1.01	81961.61
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	8.87	7.05	53.57	1.38	16.04	11.73	1.01	82119.03
最 大 值	10.91	14.38	69.21	2.34	18.01	12.46	1.06	85460.43
發生時間	5	6	5	16	2	6		2
最 小 值	4.50	5.02	36.92	0.80	14.31	11.13	0.98	75841.67
發生時間	9	22	4	9	18	2		6
零點偏移(率)	-0.05 %	0.01 %	-0.01 %	0.00 %		-0.04		0.20 %
全幅偏移(率)	0.00 %	-0.20 %	0.16 %	0.12 %		0.20		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002 樹林焚化廠CEMS系統日報表-A

報表日期: 113年02月25日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	7.80	8.40	57.36	2.43	14.65	11.80	1.01	76565.89
1	2.39	7.87	49.08	2.03	17.22	11.44	1.02	77621.02
2	2.21	7.80	50.45	1.70	16.93	11.43	1.04	77632.60
3	5.30	7.86	45.95	2.53	15.57	11.60	1.04	77975.49
4	7.96	7.75	47.51	3.31	15.11	11.59	1.02	78441.61
5	4.39	7.35	52.09	2.03	15.51	11.70	1.02	76985.77
6	1.73	6.89	57.38	1.20	16.77	11.37	1.02	74428.26
7	2.00	6.63	43.73	1.76	16.67	11.45	1.01	77818.61
8	5.66	4.10	42.82	2.35	16.85	10.89	0.99	82189.94
9	7.79	3.70	56.30	2.14	16.67	11.03	0.99	80947.54
10	7.95	3.96	58.18	2.55	15.72	11.37	0.98	79152.92
11	8.01	3.47	59.91	2.44	16.08	11.02	0.98	81901.63
12	9.00	3.85	59.93	3.45	14.52	11.51	0.99	78780.63
13	8.90	3.39	55.90	3.19	14.77	11.36	0.99	80517.86
14	9.75	3.60	56.05	3.33	14.39	11.50	0.98	79631.52
15	10.44	3.63	47.77	3.35	15.59	11.06	0.99	81675.74
16	6.55	3.59	59.31	3.11	14.62	11.62	1.02	78744.24
17	6.28	4.02	61.78	2.41	16.69	10.96	1.03	82277.94
18	10.17	4.39	54.85	2.75	15.03	11.29	1.03	80638.27
19	8.73	5.87	58.57	2.36	15.82	11.08	1.04	81792.13
20	8.66	6.42	57.59	2.73	14.48	11.51	1.02	79700.90
21	9.21	8.85	47.28	2.90	14.32	11.43	1.02	80169.12
22	7.85	3.04	59.22	2.59	14.24	11.70	1.02	78064.13
23	8.30	3.69	61.62	2.51	13.78	11.85	1.02	77358.36
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	6.96	5.42	54.19	2.55	15.50	11.40	1.01	79208.84
最 大 值	10.44	8.85	61.78	3.45	17.22	11.85	1.04	82277.94
發生時間	15	21	17	12	1	23		17
最 小 值	1.73	3.04	42.82	1.20	13.78	10.89	0.97	74428.26
發生時間	6	22	8	6	23	8		6
零點偏移(率)	-0.01 %	-0.05 %	-0.02 %	-0.01 %		-0.04		0.10 %
全幅偏移(率)	0.13 %	-0.33 %	-0.57 %	0.21 %		0.11		0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

