

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 113年06月18日

報表日期: 113年06月18日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	10.95	4.82	55.73	3.24	17.54	11.11	0.95	84946.03
1	9.05	4.86	54.24	3.78	18.12	11.08	0.95	84597.03
2	9.34	4.61	56.82	2.96	18.51	11.05	0.96	84940.71
3	9.83	4.68	68.57	1.84	20.21	10.84	0.97	84758.64
4	8.12	4.70	62.40	1.17	22.19	9.79	0.97	91422.61
5	10.90	5.00	45.22	2.48	20.46	10.30	1.01	92050.51
6	9.17	3.36	60.27	1.91	19.94	11.02	1.02	83593.93
7	8.86	3.85	54.31	1.75	21.72	10.29	1.01	87613.03
8	8.15	4.26	53.29	1.14	22.10	10.38	1.01	86707.18
9	8.56	5.28	36.82	2.22	19.60	10.89	1.02	85016.97
10	9.33	9.52	33.96	3.71	16.20	11.40	1.02	83937.25
11	11.94	12.16	48.38	4.74	16.12	11.48	1.00	82247.93
12	10.72	4.10	38.86	2.62	16.07	11.27	0.98	79791.83
13	13.06	3.90	44.87	3.35	17.23	11.23	0.99	82708.77
14	11.79	2.67	55.16	1.93	18.18	11.11	1.00	83604.72
15	11.35	1.58	54.27	3.02	18.00	11.27	0.99	85038.62
16	9.15	3.31	43.45	3.89	19.33	10.58	0.97	88655.84
17	6.28	3.02	49.39	1.94	18.64	11.07	0.98	84704.37
18	8.44	3.44	48.08	2.52	18.83	10.64	0.97	87716.15
19	9.20	3.05	53.92	3.27	20.29	10.53	0.97	86995.79
20	9.07	3.84	48.34	2.91	18.89	10.87	0.99	85362.96
21	7.73	4.90	47.18	2.82	18.32	10.94	0.97	85570.87
22	9.99	3.43	47.09	3.70	18.01	10.90	0.98	85745.55
23	8.89	4.58	52.16	2.73	17.91	11.35	0.99	82020.78
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	9.58	4.54	50.53	2.73	18.85	10.89	0.99	85406.17
最 大 值	13.06	12.16	68.57	4.74	22.19	11.48	1.02	92050.51
發生時間	13	11	3	11	4	11		5
最 小 值	6.28	1.58	33.96	1.14	16.07	9.79	0.94	79791.83
發生時間	17	15	10	8	12	4		12
零點偏移(率)	-0.02 %	-0.00 %	-0.19 %	-0.01 %		-0.07		0.00 %
全幅偏移(率)	0.00 %	-0.34 %	0.73 %	-0.20 %		0.12		0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 113年06月18日

項目 單位 小時\限標	#3-HCl_C ppm	#3-CO_C ppm	#3-NOx_C ppm	#3-SO2_C ppm	#3-H2O %	#3-O2_C %	#3-OPAC %	#3-FLOW_C m3/Hour
0	8.35	2.72	60.24	0.96	19.98	10.04	0.96	82888.13
1	4.93	3.29	47.92	0.69	20.90	9.68	0.94	84843.17
2	3.11	3.43	60.32	0.85	17.70	11.06	0.93	76477.65
3	5.00	2.77	56.55	1.02	19.90	9.69	0.98	86099.03
4	4.91	3.49	56.40	0.91	19.07	10.26	0.98	82368.78
5	11.35	2.53	52.00	1.21	18.41	10.33	0.96	82980.00
6	11.89	2.72	55.31	1.69	17.98	10.54	0.97	81152.28
7	7.72	3.23	55.18	1.42	17.97	10.53	0.98	80510.05
8	11.29	3.45	50.23	1.75	17.76	10.40	0.96	82168.19
9	6.06	4.45	55.30	1.21	16.84	10.76	0.95	79697.80
10	11.65	2.74	61.61	1.83	17.36	10.77	0.99	78971.21
11	6.85	4.95	46.67	1.81	16.14	11.13	0.99	77864.59
12	4.68	2.59	54.45	1.01	16.99	10.78	0.98	78561.21
13	5.51	4.16	55.85	1.11	18.59	10.31	0.98	81704.20
14	11.96	5.56	50.50	1.66	18.85	10.25	0.96	81175.80
15	2.81	2.84	55.11	1.26	17.76	10.55	0.96	78447.30
16	3.17	3.36	50.77	1.94	18.11	10.20	0.94	81979.14
17	1.18	4.58	48.43	0.72	18.73	10.06	0.93	83887.48
18	2.74	1.57	54.91	1.78	18.56	10.49	0.94	80273.43
19	6.83	1.77	50.26	0.77	19.74	9.67	0.94	86052.04
20	7.89	2.94	51.48	0.72	18.41	10.05	0.97	84471.59
21	4.43	2.08	55.89	0.64	18.15	10.19	0.98	83861.53
22	3.23	2.68	54.89	0.91	17.57	10.67	0.97	80500.66
23	10.60	2.52	61.46	1.09	17.64	10.86	0.97	79448.45
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	6.59	3.18	54.24	1.21	18.30	10.39	0.96	81515.99
最 大 值	11.96	5.56	61.61	1.94	20.90	11.13	1.01	86099.03
發生時間	14	14	10	16	1	11		3
最 小 值	1.18	1.57	46.67	0.64	16.14	9.67	0.92	76477.65
發生時間	17	18	11	21	11	19		2
零點偏移(率)	-0.06 %	0.00 %	-0.03 %	0.00 %		-0.08		0.20 %
全幅偏移(率)	0.02 %	-0.00 %	-0.17 %	-0.20 %		0.31		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								