

## P001\_樹林焚化廠CEMS系統日報表-A

報表日期: 115年03月28日

項目	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	9.43	13.09	46.89	0.98	18.05	11.45	1.35	78482.20
1	9.09	12.65	54.26	1.22	17.75	11.15	1.35	80549.31
2	9.60	12.38	52.84	1.67	19.06	10.75	1.34	82063.76
3	9.29	12.39	50.65	1.72	17.54	10.87	1.34	81406.77
4	10.26	12.74	48.76	1.83	17.17	11.01	1.34	80401.74
5	7.80	11.03	47.60	1.74	18.83	10.82	1.34	84725.69
6	9.91	11.32	40.92	3.04	16.57	11.21	1.34	81403.28
7	9.94	11.25	57.49	2.47	17.40	11.11	1.35	82318.29
8	9.05	11.50	58.05	3.31	15.84	11.40	1.34	79368.12
9	10.06	11.94	60.82	3.73	15.32	11.50	1.34	78546.09
10	9.47	11.46	58.43	2.91	16.20	11.29	1.34	79036.77
11	12.96	11.73	54.95	2.39	17.24	11.30	1.34	78120.28
12	7.35	11.27	46.09	3.23	16.39	11.12	1.34	77387.72
13	8.85	12.91	57.82	3.53	16.03	11.20	1.34	76944.16
14	10.51	13.65	56.97	3.43	15.36	11.31	1.34	76312.40
15	8.71	13.32	54.37	3.03	16.57	11.33	1.34	76665.93
16	9.35	12.85	49.44	2.36	17.82	10.76	1.34	80145.86
17	9.04	13.37	55.84	1.71	18.41	11.23	1.34	76486.55
18	9.22	13.69	45.38	2.16	17.90	10.90	1.34	82942.42
19	9.32	12.75	62.42	1.37	19.50	10.97	1.34	80414.24
20	10.23	12.30	48.31	2.10	18.29	10.64	1.34	82259.26
21	8.55	12.47	51.19	1.23	20.10	10.59	1.34	82456.36
22	9.17	12.60	45.39	2.85	18.36	10.56	1.34	84815.32
23	11.32	12.88	53.57	1.87	19.17	11.08	1.34	78244.07
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	9.52	12.40	52.44	2.33	17.54	11.06	1.34	80062.36
最大值	12.96	13.69	62.42	3.73	20.10	11.50	1.36	84815.32
發生時間	11	18	19	9	21	9		22
最小值	7.35	11.03	40.92	0.98	15.32	10.56	1.33	76312.40
發生時間	12	5	6	0	9	22		14
零點偏移(率)	0.00 %	-0.00 %	-0.14 %	-0.01 %		-0.06	-0.01	0.30 %
全幅偏移(率)	0.62 %	-0.82 %	-0.36 %	-0.40 %		0.35	0.10	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002\_樹林焚化廠CEMS系統日報表-A

報表日期: 115年03月28日

項目	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	10.20	10.09	49.29	0.52	17.79	11.02	0.83	75680.46
1	12.52	9.49	59.68	1.09	19.12	10.54	0.83	82448.51
2	10.08	9.14	48.44	0.83	18.58	10.32	0.83	81315.99
3	12.13	8.63	57.64	0.79	18.22	10.68	0.83	81784.40
4	10.60	11.53	64.76	0.58	18.91	10.44	0.83	80554.16
5	11.29	10.17	43.33	1.01	17.41	10.89	0.83	76198.52
6	10.65	7.86	50.78	1.18	17.29	10.79	0.84	79141.45
7	10.11	9.76	61.82	1.83	16.40	11.12	0.84	80291.37
8	16.98	10.14	65.59	1.81	17.24	10.92	0.84	80682.98
9	14.08	9.44	62.43	0.90	18.04	10.79	0.84	82742.83
10	9.34	8.07	60.08	1.02	17.55	11.05	0.83	79479.73
11	10.90	9.02	46.07	0.76	17.88	10.69	0.83	80167.54
12	9.73	10.09	58.64	0.70	17.67	10.98	0.83	77314.08
13	9.36	9.72	55.80	0.65	18.49	10.84	0.83	79138.06
14	10.44	9.34	58.13	0.92	17.65	11.03	0.84	80185.35
15	8.34	9.22	57.01	0.71	18.11	11.21	0.84	77875.42
16	10.51	10.84	53.29	0.92	18.39	10.98	0.84	78648.36
17	10.08	13.00	52.88	1.44	17.01	11.35	0.83	77408.40
18	11.64	10.13	57.87	1.20	17.43	11.39	0.83	75166.68
19	9.33	12.58	62.31	0.63	19.18	10.90	0.83	77916.40
20	10.64	19.91	60.54	0.97	17.61	11.62	0.83	66915.76
21	10.55	9.50	53.86	0.99	17.08	11.35	0.83	71487.66
22	8.89	9.03	56.41	0.88	17.35	11.35	0.83	74184.44
23	9.39	9.67	50.58	1.06	17.10	11.36	0.84	74215.18
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	10.74	10.27	56.13	0.97	17.81	10.98	0.83	77955.99
最大值	16.98	19.91	65.59	1.83	19.18	11.62	0.87	82742.83
發生時間	8	20	8	7	19	20		9
最小值	8.34	7.86	43.33	0.52	16.40	10.32	0.83	66915.76
發生時間	15	6	5	0	7	2		20
零點偏移(率)	-0.08 %	0.00 %	-0.06 %	-0.01 %		-0.10	-0.06	-0.10 %
全幅偏移(率)	-2.53 %	0.96 %	0.03 %	-0.00 %		-0.09	0.31	-0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

## P003\_樹林焚化廠CEMS系統日報表-A

報表日期: 115年03月28日

項目	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
單位	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
小時\限標	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.60	11.09	55.18	2.97	18.34	9.74	1.34	75239.92
1	9.08	11.23	59.59	2.83	18.86	9.80	1.33	76149.23
2	11.61	11.52	56.04	3.21	18.00	10.03	1.34	76252.59
3	11.30	11.75	53.62	3.40	15.76	10.56	1.32	73342.42
4	10.40	11.86	56.49	3.15	16.82	10.29	1.32	76809.40
5	8.92	12.29	54.23	3.38	15.35	10.72	1.33	72965.10
6	9.53	11.87	53.45	3.32	14.92	10.63	1.33	72050.88
7	11.92	10.60	46.82	3.53	15.95	10.58	1.34	72152.13
8	9.07	10.63	43.53	3.13	16.34	10.20	1.34	74106.89
9	10.86	10.76	45.52	3.15	16.99	9.90	1.35	75656.10
10	9.42	10.54	49.86	2.98	17.10	10.23	1.36	76205.39
11	8.46	10.85	49.40	2.94	16.14	10.42	1.35	75290.40
12	9.06	11.66	51.22	3.28	15.13	10.58	1.33	71816.20
13	9.44	12.81	51.96	3.60	14.29	10.60	1.34	72088.22
14	10.07	12.85	48.56	3.26	15.71	10.19	1.34	71625.10
15	10.28	11.30	54.11	3.26	17.02	9.85	1.34	74271.20
16	9.16	10.92	54.31	3.05	17.88	9.73	1.31	75674.16
17	7.88	10.56	54.68	2.78	18.03	9.75	1.31	75582.49
18	9.61	10.61	58.83	2.92	18.32	9.78	1.33	75957.13
19	8.27	10.80	56.78	2.76	18.51	9.94	1.33	75078.11
20	9.14	10.67	55.89	2.85	18.18	9.83	1.34	74864.84
21	9.77	10.45	58.11	2.87	18.25	9.59	1.31	75691.45
22	8.97	10.86	53.14	2.93	19.27	9.34	1.33	77159.06
23	9.57	11.02	50.71	2.94	19.19	9.51	1.36	74874.46
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平均值	9.60	11.23	53.00	3.10	17.10	10.07	1.33	74620.95
最大值	11.92	12.85	59.59	3.60	19.27	10.72	1.37	77159.06
發生時間	7	14	1	13	22	5		22
最小值	7.88	10.45	43.53	2.76	14.29	9.34	1.31	71625.10
發生時間	17	21	8	19	13	22		14
零點偏移(率)	-0.08 %	0.00 %	-0.18 %	-0.03 %		-0.09	-0.01	0.20 %
全幅偏移(率)	0.38 %	0.13 %	0.02 %	0.10 %		0.26	0.04	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								