

P001 樹林焚化廠CEMS系統日報表-A

報表日期: 114年07月28日

項目 單位 小時\限標	#1-HCl_C ppm	#1-CO_C ppm	#1-NOx_C ppm	#1-SO2_C ppm	#1-H2O %	#1-O2_C %	#1-OPAC %	#1-FLOW_C m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	9.68	8.87	57.90	3.57	17.25	11.47	1.31	80713.98
1	6.04	9.21	69.92	2.29	17.95	11.77	1.31	78150.86
2	7.32	8.92	49.41	4.05	16.74	11.60	1.33	80510.11
3	1.33	9.57	58.39	3.31	15.35	12.58	1.33	72421.05
4	3.48	9.46	55.79	2.27	16.69	12.02	1.31	76912.25
5	6.57	9.13	55.31	3.03	15.87	11.89	1.30	77528.64
6	6.08	9.84	50.95	3.76	16.41	11.70	1.31	79873.01
7	8.83	9.17	50.80	2.23	15.15	12.21	1.31	76767.41
8	8.62	8.68	57.12	1.61	16.20	12.13	1.32	75905.20
9	9.17	8.65	49.04	2.12	15.46	11.97	1.33	78682.46
10	8.53	9.34	53.26	2.02	15.56	12.19	1.33	76449.44
11	9.38	9.23	47.76	2.42	14.95	12.17	1.30	76795.12
12	6.26	9.15	47.44	1.95	15.43	12.25	1.33	75584.55
13	8.51	8.91	56.45	1.74	15.57	12.16	1.32	76103.00
14	9.22	11.67	51.86	1.97	14.91	12.29	1.30	75720.73
15	9.85	10.70	50.84	2.79	13.45	12.66	1.30	73587.67
16	8.03	10.01	54.21	2.57	15.10	12.33	1.31	75185.28
17	7.26	10.97	57.69	1.75	15.66	12.06	1.31	76680.23
18	7.53	9.28	52.26	1.84	15.13	12.07	1.31	77392.19
19	7.18	9.48	56.72	1.44	16.17	12.18	1.30	76024.48
20	6.51	9.30	55.93	1.14	16.62	11.89	1.30	78027.82
21	5.96	9.43	57.86	1.04	16.73	11.97	1.31	77273.78
22	0.45	9.46	65.45	0.98	16.54	11.95	1.32	77997.33
23	0.33	10.34	51.94	2.21	15.82	12.04	1.32	77186.19
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	6.75	9.53	54.76	2.25	15.86	12.06	1.31	76978.03
最 大 值	9.85	11.67	69.92	4.05	17.95	12.66	1.34	80713.98
發生時間	15	14	1	2	1	15		0
最 小 值	0.33	8.65	47.44	0.98	13.45	11.47	1.29	72421.05
發生時間	23	9	12	22	15	0		3
零點偏移(率)	-0.28 %	-0.00 %	-0.03 %	-0.01 %		-0.19		0.10 %
全幅偏移(率)	-0.20 %	0.51 %	-0.55 %	0.33 %		-0.24		0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A 報表日期: 114年07月28日

報表日期: 114年07月28日

項目 單位 小時\限標	#3-HCl_C ppm	#3-CO_C ppm	#3-NOx_C ppm	#3-SO2_C ppm	#3-H2O %	#3-O2_C %	#3-OPAC %	#3-FLOW_C m3/Hour
0	2.99	10.79	57.94	2.90	17.46	9.35	1.41	76605.78
1	0.72	9.11	64.70	2.62	17.86	9.51	1.38	75875.58
2	5.39	7.02	63.08	3.04	18.61	9.65	1.38	77351.66
3	9.31	7.49	48.77	3.16	18.03	9.22	1.37	77985.64
4	6.39	8.58	54.19	3.76	18.24	9.20	1.38	78476.61
5	1.35	7.51	55.57	3.32	18.39	9.14	1.38	82072.53
6	0.92	8.19	49.32	3.03	17.64	9.52	1.38	76040.51
7	3.30	10.58	53.84	2.46	17.13	9.43	1.42	76486.58
8	1.73	10.44	48.43	2.29	18.24	9.00	1.45	76075.01
9	2.97	11.34	46.11	2.41	17.82	9.12	1.47	77749.06
10	1.64	9.30	57.54	2.28	18.62	8.87	1.48	75604.56
11	1.04	11.92	47.41	1.99	18.02	8.71	1.43	73554.66
12	2.91	13.50	53.60	2.10	18.56	8.76	1.45	75232.27
13	2.04	9.54	56.61	1.95	18.08	8.96	1.48	76740.95
14	3.26	9.63	58.50	2.05	17.57	9.11	1.48	76809.57
15	4.61	9.92	65.52	2.30	18.73	9.20	1.45	75510.83
16	6.65	7.71	48.93	1.80	18.05	9.27	1.43	83199.91
17	2.71	9.52	45.03	1.09	18.04	9.29	1.43	80883.22
18	2.59	8.75	54.63	1.13	18.89	9.38	1.43	81291.75
19	3.49	8.26	49.08	1.19	19.18	9.18	1.43	82117.54
20	2.59	7.79	54.45	1.16	18.51	9.71	1.47	81439.15
21	9.09	8.29	50.76	1.32	16.84	9.91	1.45	83356.30
22	11.14	8.12	53.66	1.85	16.99	10.18	1.43	78297.24
23	2.61	8.65	57.00	2.36	18.82	9.01	1.43	78112.00
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	3.81	9.25	53.94	2.23	18.10	9.28	1.43	78202.87
最 大 值	11.14	13.50	65.52	3.76	19.18	10.18	1.51	83356.30
發生時間	22	12	15	4	19	22		21
最 小 值	0.72	7.02	45.03	1.09	16.84	8.71	1.37	73554.66
發生時間	1	2	17	17	21	11		11
零點偏移(率)	-0.01 %	-0.00 %	-0.01 %	-0.04 %		-0.06		-0.10 %
全幅偏移(率)	-1.04 %	0.13 %	-0.58 %	-0.00 %		-0.09		-0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								