

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 115年07月01日

報表日期: 115年07月01日

項目 單位 小時\限標	#1-HCl_C ppm	#1-CO_C ppm	#1-NOx_C ppm	#1-SO2_C ppm	#1-H2O %	#1-O2_C %	#1-OPAC %	#1-FLOW_C m3/Hour
0	14.72	11.04	51.58	2.56	19.75	10.54	1.19	86097.55
1	3.75	11.74	49.72	2.23	19.71	11.16	1.19	80662.90
2	14.73	12.51	43.33	4.41	18.82	10.67	1.19	85564.39
3	2.15	10.95	44.60	3.20	19.56	10.35	1.19	87701.37
4	3.36	11.38	42.20	3.58	18.35	10.84	1.18	84503.86
5	6.76	11.34	52.81	2.13	18.74	10.64	1.20	89159.21
6	4.14	11.60	55.13	1.60	18.55	11.09	1.22	82431.23
7	11.80	11.48	55.87	2.18	18.29	11.07	1.22	82750.59
8	9.01	11.16	53.36	2.59	18.50	10.85	1.20	84478.06
9	11.20	11.33	51.17	3.65	17.35	11.15	1.21	82824.69
10	8.63	11.75	44.61	3.11	17.02	10.84	1.24	85184.34
11	10.18	12.83	43.65	2.90	16.84	11.09	1.23	83773.78
12	15.67	11.35	46.07	4.47	15.95	11.22	1.22	83732.25
13	4.66	11.88	45.51	4.40	15.83	11.60	1.22	80734.35
14	6.57	12.28	43.60	4.69	15.24	12.01	1.20	78517.98
15	6.06	12.22	42.26	3.51	15.01	12.06	1.19	77662.42
16	10.62	12.01	60.79	2.68	15.59	11.80	1.19	78528.05
17	8.18	11.95	60.39	2.08	18.24	10.74	1.20	83675.99
18	8.37	11.29	57.08	2.62	17.51	10.87	1.20	84206.17
19	8.33	11.52	54.64	1.68	17.09	11.15	1.20	82367.67
20	7.73	13.17	49.01	1.83	16.97	11.15	1.21	82314.85
21	9.78	13.16	56.63	1.85	16.83	11.46	1.24	79793.88
22	8.11	15.06	59.85	2.32	15.70	12.89	1.24	68947.96
23	6.66	14.03	56.16	2.61	15.64	12.97	1.25	68083.75
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	8.38	12.04	50.83	2.87	17.38	11.26	1.21	81820.72
最 大 值	15.67	15.06	60.79	4.69	19.75	12.97	1.28	89159.21
發生時間	12	22	16	14	0	23		5
最 小 值	2.15	10.95	42.20	1.60	15.01	10.35	1.18	68083.75
發生時間	3	3	4	6	15	3		23
零點偏移(率)	-0.01 %	0.01 %	-0.06 %	-0.03 %		-0.03	-0.01	0.20 %
全幅偏移(率)	1.75 %	0.93 %	-0.80 %	1.68 %		0.01	-0.06	0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A 報表日期: 115年07月01日

報表日期: 115年07月01日

項目 單位 小時\限標	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	2.86	11.49	61.05	3.00	21.15	9.43	1.31	70535.06
1	6.99	12.23	40.52	2.54	22.14	8.58	1.30	79330.03
2	6.23	11.84	53.82	3.34	19.25	9.12	1.29	75581.67
3	6.86	10.79	59.42	3.48	19.40	9.10	1.27	77674.84
4	4.56	11.12	53.73	3.20	19.33	9.54	1.29	74157.04
5	6.46	12.08	45.88	3.48	20.27	9.17	1.30	75608.33
6	2.05	11.00	52.14	1.50	19.71	9.48	1.29	74896.73
7	3.20	10.84	60.95	4.65	19.16	9.24	1.31	78044.02
8	3.64	11.02	63.47	2.26	18.18	9.68	1.28	73949.36
9	5.68	11.37	58.37	2.29	18.31	9.30	1.31	74165.16
10	9.79	9.15	49.05	3.00	18.89	8.30	1.28	76906.95
11	8.63	9.06	54.74	1.97	19.20	8.76	1.27	74550.10
12	7.34	10.56	47.43	2.00	18.43	8.38	1.29	73567.92
13	4.85	10.32	47.90	1.89	18.10	8.49	1.28	71857.01
14	6.98	10.73	47.57	2.01	17.28	8.85	1.26	71314.06
15	3.88	10.34	53.47	1.90	17.60	8.85	1.25	70361.53
16	2.64	11.24	50.41	3.10	16.60	8.76	1.25	72326.02
17	4.24	11.37	55.94	3.60	16.41	9.28	1.26	71324.55
18	11.79	10.94	61.38	2.97	17.51	9.28	1.30	76846.63
19	9.63	11.21	62.94	2.90	17.61	9.71	1.30	75133.84
20	6.83	12.95	57.47	2.99	19.36	9.62	1.28	73100.39
21	12.58	12.06	56.21	2.95	19.08	9.16	1.27	77619.26
22	9.62	12.30	63.88	3.29	19.81	9.08	1.27	77598.91
23	10.33	12.12	52.92	3.06	18.98	8.67	1.27	79047.09
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	6.57	11.17	54.61	2.81	18.82	9.08	1.28	74812.35
最 大 值	12.58	12.95	63.88	4.65	22.14	9.71	1.32	79330.03
發生時間	21	20	22	7	1	19		1
最 小 值	2.05	9.06	40.52	1.50	16.41	8.30	1.24	70361.53
發生時間	6	11	1	6	17	10		15
零點偏移(率)	-0.08 %	-0.14 %	-0.04 %	-0.11 %		-0.02	-0.01	-0.10 %
全幅偏移(率)	-0.70 %	-0.71 %	-1.06 %	-0.85 %		0.33	0.00	-0.10 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								