

P001 樹林焚化廠CEMS系統日報表-A 報表日期: 113年08月27日

報表日期: 113年08月27日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	4.81	2.82	48.68	1.75	17.09	11.67	0.98	80583.74
1	4.78	2.55	50.89	1.56	17.79	11.48	0.97	81631.28
2	4.41	2.61	55.41	1.44	18.46	11.48	0.97	81106.61
3	5.37	2.65	50.18	1.57	17.93	11.28	0.94	83216.89
4	5.76	2.74	61.54	1.27	18.58	11.26	0.96	82856.00
5	6.60	3.51	55.05	1.45	19.37	10.81	0.97	88168.67
6	6.16	3.33	47.19	1.04	19.53	10.90	0.97	84709.12
7	4.69	3.31	51.96	1.43	17.94	11.38	0.97	82169.75
8	5.46	3.82	58.11	0.81	20.06	10.74	0.98	85683.89
9	4.83	3.63	56.45	1.72	17.34	12.15	0.97	77216.67
10	6.08	3.76	49.46	2.77	15.70	13.13	0.95	76249.46
11	6.93	3.30	50.77	2.18	15.34	13.42	0.97	75364.09
12	6.65	3.83	55.12	1.77	15.80	13.35	0.94	76044.62
13	11.02	4.19	46.16	2.31	15.58	12.16	0.94	78710.62
14	6.74	3.87	49.75	1.20	17.48	11.94	0.95	77371.43
15	4.59	3.52	65.63	0.50	18.53	11.88	0.97	77554.32
16	1.06	3.59	60.97	0.34	19.97	11.37	0.98	80476.57
17	3.02	8.26	38.38	0.75	20.04	10.81	0.96	85098.06
18	5.16	6.60	39.47	2.02	17.43	11.36	0.97	82859.20
19	7.17	2.89	42.65	3.11	14.79	12.02	0.97	79713.13
20	7.88	3.07	40.91	1.65	15.25	11.83	0.97	79199.54
21	7.47	1.97	39.77	1.19	15.56	11.87	0.97	80113.86
22	6.15	2.97	45.26	2.01	16.30	12.01	0.97	77774.77
23	4.54	3.56	48.28	1.47	16.79	11.91	0.97	78985.37
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	5.72	3.60	50.34	1.55	17.44	11.76	0.96	80535.74
最 大 值	11.02	8.26	65.63	3.11	20.06	13.42	1.00	88168.67
發生時間	13	17	15	19	8	11		5
最 小 值	1.06	1.97	38.38	0.34	14.79	10.74	0.92	75364.09
發生時間	16	21	17	16	19	8		11
零點偏移(率)	-0.02 %	0.00 %	-0.07 %	0.00 %		-0.06		0.00 %
全幅偏移(率)	0.00 %	0.03 %	-0.01 %	0.04 %		-0.08		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002 樹林焚化廠CEMS系統日報表-A

報表日期: 113年08月27日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	5.03	1.78	46.05	1.79	17.03	11.01	0.86	80950.90
1	4.78	1.82	49.59	1.77	18.00	11.02	0.88	80425.23
2	1.23	1.92	53.07	1.97	17.73	11.10	0.89	80043.98
3	5.70	1.82	50.48	1.63	18.19	10.79	0.90	81987.09
4	3.31	1.92	57.36	1.81	17.94	11.24	0.90	78362.92
5	2.66	2.09	58.38	1.60	18.87	10.60	0.90	82175.75
6	1.60	3.66	52.48	1.52	18.18	10.72	0.88	80788.22
7	1.85	4.01	52.91	2.32	17.81	10.72	0.87	83462.91
8	1.35	6.10	55.20	2.54	17.26	10.89	0.86	82528.85
9	10.31	2.52	54.48	2.55	17.71	11.16	0.87	82191.85
10	8.78	2.10	55.44	2.38	17.84	11.87	0.83	83107.51
11	7.56	1.25	53.98	1.93	17.19	12.07	0.82	83223.48
12	4.43	1.66	45.91	1.56	17.63	12.02	0.82	83969.59
13	11.32	1.96	38.89	2.89	15.63	11.52	0.83	83328.68
14	11.35	3.82	39.75	3.36	15.78	11.38	0.85	81791.25
15	12.01	2.05	39.51	1.56	17.48	11.18	0.88	80395.90
16	5.67	2.50	57.67	1.19	18.12	11.55	0.87	76896.48
17	11.89	7.61	38.24	2.18	17.06	11.07	0.84	83063.55
18	11.24	8.02	35.07	1.52	15.79	10.91	0.84	85384.66
19	6.15	7.18	41.68	0.88	15.90	11.06	0.87	82926.34
20	9.16	2.52	43.78	1.50	16.61	10.91	0.85	83168.29
21	9.96	3.07	43.54	1.95	16.81	10.72	0.84	85355.31
22	6.72	1.70	60.03	2.25	17.66	10.83	0.83	82467.39
23	8.95	4.55	42.90	2.43	16.11	11.21	0.83	80328.53
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	6.79	3.23	48.60	1.96	17.26	11.15	0.86	82013.53
最 大 值	12.01	8.02	60.03	3.36	18.87	12.07	0.99	85384.66
發生時間	15	18	22	14	5	11		18
最 小 值	1.23	1.25	35.07	0.88	15.63	10.60	0.81	76896.48
發生時間	2	11	18	19	13	5		16
零點偏移(率)	-0.07 %	-0.00 %	-0.01 %	0.00 %		-0.04		-0.10 %
全幅偏移(率)	0.13 %	0.05 %	-0.20 %	0.18 %		-0.29		0.00 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 113年08月27日

項目 單位 小時\限標	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.76	2.09	60.78	0.42	16.38	11.31	1.64	75233.62
1	6.58	2.24	53.81	0.39	17.09	11.04	1.68	77963.30
2	4.29	2.33	56.36	0.53	16.79	11.36	1.69	75651.46
3	12.64	2.03	59.66	1.09	15.83	11.41	1.69	77481.44
4	7.48	2.05	59.54	0.95	17.16	11.01	1.69	79346.21
5	5.27	2.34	47.87	0.44	17.63	11.04	1.69	77841.87
6	4.43	2.66	61.04	0.37	17.61	10.95	1.66	78717.92
7	3.58	3.74	43.53	0.18	16.89	11.08	1.70	76930.19
8	8.48	3.04	58.27	0.56	17.34	10.82	1.72	80697.68
9	6.74	5.94	59.46	0.51	17.88	10.83	1.69	80218.11
10	7.18	2.54	60.61	3.04	17.52	11.87	1.69	78981.66
11	5.17	4.01	62.20	2.01	18.15	12.20	1.72	73750.88
12	5.02	2.03	47.16	2.88	18.34	11.36	1.69	83497.81
13	10.20	2.57	37.35	2.63	15.16	11.26	1.66	78632.61
14	7.54	2.91	43.82	1.88	16.63	11.21	1.67	77708.64
15	9.11	2.48	57.21	1.98	18.78	10.71	1.69	79292.56
16	6.67	3.55	40.31	1.83	18.97	10.33	1.69	82228.56
17	6.53	1.82	55.29	1.84	18.91	10.40	1.68	81947.02
18	5.51	1.59	56.95	1.84	18.49	10.45	1.69	81857.36
19	10.69	2.36	54.53	2.00	16.60	10.89	1.70	80288.85
20	3.34	1.43	56.65	1.66	17.33	10.69	1.68	80990.62
21	9.51	3.62	48.04	1.96	16.48	10.86	1.67	80148.92
22	8.23	3.46	53.71	2.23	15.90	11.49	1.70	75932.38
23	8.94	2.47	54.40	1.71	17.75	10.93	1.71	78791.19
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	7.16	2.72	53.69	1.46	17.32	11.06	1.69	78922.12
最 大 值	12.64	5.94	62.20	3.04	18.97	12.20	1.80	83497.81
發生時間	3	9	11	10	16	11		12
最 小 值	3.34	1.43	37.35	0.18	15.16	10.33	1.62	73750.88
發生時間	20	20	13	7	13	16		11
零點偏移(率)	-0.04 %	-0.07 %	-0.09 %	-0.01 %		-0.09		0.00 %
全幅偏移(率)	0.02 %	-0.19 %	-0.01 %	0.31 %		-0.14		0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								